



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110712

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/27/2020	12:25
2	CAL2	1	Water	NA	NA	20.1	7.01	08/27/2020	12:26
3	CAL3	1	Water	NA	NA	20.1	10.03	08/27/2020	12:27
4	ICV	1	Water	NA	NA	20.1	6.98	08/27/2020	12:29
5	CCV1	1	Water	NA	NA	20.2	2.01	08/27/2020	12:30
6	L3822-01	1	Water	NA	NA	23.0	5.75	08/27/2020	12:32
7	L3822-01DUP	1	Water	NA	NA	23.1	5.76	08/27/2020	12:33
8	CCV2	1	Water	NA	NA	20.1	12.02	08/27/2020	12:35

LB 110712

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3822 WorkList ID : 141821 Department : Wet-Chemistry Date : 08-27-2020 12:17:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2020	Water	L3822-01	pH	Cool 4 deg C	STER08	L61	POOL-WATER	08/21/2020	SM 4500-H B

Date/Time 08/27/2020 12:00
 Received by: Jy (wet-chem)
 Relinquished by: AM (wet-chem)

Date/Time 08/27/2020 13:00
 Received by: AM (wet-chem)
 Relinquished by: Jy (wet-chem)