


7471B

LB110719
INSTRUMENT ID: CV4

Linear

 μ Abs.:

46940



A= 0.0000e+000

B= 2.1324e-004

C= 3.9026e-002

Rho= 0.9996148

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.050	0.050	50	0.000	50				
0.2	0.200	0.244	0.044	959	0.0 %	959				
2.5	2.500	2.333	-0.167	10759	0.0 %	10759				
5.0	5.000	5.131	0.131	23879	0.0 %	23879				
7.5	7.500	7.394	-0.106	34491	0.0 %	34491				
10.0	10.000	10.049	0.049	46940	0.0 %	46940				

LB110719 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 27 Aug 2020 14:36:14					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	50	-	0.0000	7471B	PPB	27 Aug 2020 14:41:29	S	Std
0.2 - 1	50.2	959	-	0.2000	7471B	PPB	27 Aug 2020 14:43:47	S	Std
2.5 - 1	52.5	10759	-	2.5000	7471B	PPB	27 Aug 2020 14:46:04	S	Std
5.0 - 1	55	23879	-	5.0000	7471B	PPB	27 Aug 2020 14:48:21	S	Std
7.5 - 1	59.5	34491	-	7.5000	7471B	PPB	27 Aug 2020 14:53:17	S	Std
10.0 - 1	510	46940	-	10.0000	7471B	PPB	27 Aug 2020 14:55:33	S	Std
ICV87 - 1	ICV87	18326	3.9469	-	7471B	PPB	27 Aug 2020 14:58:39	U	SMPL
ICB87 - 1	ICB87	-98	0.0181	-	7471B	PPB	27 Aug 2020 15:00:54	U	SMPL
CCV92 - 1	CCV92	24308	5.2225	-	7471B	PPB	27 Aug 2020 15:03:11	U	SMPL
CCB92 - 1	CCB92	-116	0.0143	-	7471B	PPB	27 Aug 2020 15:05:26	U	SMPL
CRA - 1	CRA	937	0.2388	-	7471B	PPB	27 Aug 2020 15:07:42	U	SMPL
HighStd - 1	HighStd	46645	9.9857	-	7471B	PPB	27 Aug 2020 15:09:58	U	SMPL
ChkStd - 1	ChkStd	34814	7.4628	-	7471B	PPB	27 Aug 2020 15:12:13	U	SMPL
PB131237BL - 1	PBS	-285	-0.0217	-	7471B	PPB	27 Aug 2020 15:14:31	U	SMPL
PB131237BS - 1	LCSS	17587	3.7893	-	7471B	PPB	27 Aug 2020 15:16:49	U	SMPL
L3498-01 - 1	HR-06-082620	18647	4.0153	-	7471B	PPB	27 Aug 2020 15:19:06	U	SMPL
L3498-01DUP - 1	HR-06-082620DUP	16174	3.4880	-	7471B	PPB	27 Aug 2020 15:23:59	U	SMPL
L3498-01MS - 1	HR-06-082620MS	35723	7.6566	-	7471B	PPB	27 Aug 2020 15:36:06	U	SMPL
L3498-01MSD - 1	HR-06-082620MSD	35813	7.6758	-	7471B	PPB	27 Aug 2020 15:38:23	U	SMPL
L3806-03 - 1	TP-14A-S	2005	0.4666	-	7471B	PPB	27 Aug 2020 15:40:40	U	SMPL
CCV93 - 1	CCV93	24596	5.2839	-	7471B	PPB	27 Aug 2020 15:42:57	U	SMPL
CCB93 - 1	CCB93	-57	0.0269	-	7471B	PPB	27 Aug 2020 15:45:13	U	SMPL
L3806-08 - 1	TP-14-D	1193	0.2934	-	7471B	PPB	27 Aug 2020 15:47:30	U	SMPL
L3498-01LX5 - 1		4511	1.0010	-	7471B	PPB	27 Aug 2020 15:49:45	U	SMPL
L3498-01A - 1		35308	7.5681	-	7471B	PPB	27 Aug 2020 15:52:00	U	SMPL
CCV94 - 1	CCV94	24546	5.2732	-	7471B	PPB	27 Aug 2020 15:54:16	U	SMPL
CCB94 - 1	CCB94	-61	0.0260	-	7471B	PPB	27 Aug 2020 15:56:33	U	SMPL