



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110725

Slope : 97.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	08/27/2020	16:25
2	CAL2	1	Water	NA	NA	20.2	7.00	08/27/2020	16:26
3	CAL3	1	Water	NA	NA	20.2	10.03	08/27/2020	16:27
4	ICV	1	Water	NA	NA	20.2	7.02	08/27/2020	16:30
5	CCV1	1	Water	NA	NA	20.3	2.01	08/27/2020	16:31
6	L3507-17	1	Solid	20.03	20	25.9	8.74	08/27/2020	16:33
7	L3507-17DUP	1	Solid	20.04	20	25.8	8.75	08/27/2020	16:34
8	CCV2	1	Water	NA	NA	20.2	12.02	08/27/2020	16:35

VB 110725

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3823 WorkList ID : 141833 Department : Wet-Chemistry Date : 08-27-2020 15:28:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/27/2020	Solid	L-3507-17	Corrosivity	Cool 4 deg C	PSEG05		SU-03-082720	08/27/2020	9045D

Date/Time 08/27/2020 15:30
Received by: JH 1 wet (chem)
Relinquished by: KS (wet-chem)

Date/Time 08/27/2020 17:10
Received by: KS (wet-chem)
Relinquished by: JH (wet-chem)