


LB110729

7471B

INSTRUMENT ID: CV1

 μ Abs.:

49313



Linear

A= 0.0000e+000

B= 2.0251e-004

C= 2.3534e-002

Rho= 0.9999568

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	15	0.000	15				
0.2	0.200	0.210	0.010	921	0.0 %	921				
2.5	2.500	2.433	-0.067	11899	0.0 %	11899				
5.0	5.000	5.037	0.037	24758	0.0 %	24758				
7.5	7.500	7.483	-0.017	36834	0.0 %	36834				
10.0	10.000	10.010	0.010	49313	0.0 %	49313				

LB110729 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Aug 2020 08:58:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.0000	7471B	PPB	28 Aug 2020 10:29:21	S	Std
0.2 - 1	50.2	921	-	0.2000	7471B	PPB	28 Aug 2020 10:31:38	S	Std
2.5 - 1	52.5	11899	-	2.5000	7471B	PPB	28 Aug 2020 10:33:55	S	Std
5.0 - 1	55	24758	-	5.0000	7471B	PPB	28 Aug 2020 10:36:12	S	Std
7.5 - 1	57.5	36834	-	7.5000	7471B	PPB	28 Aug 2020 10:38:29	S	Std
10.0 - 1	510	49313	-	10.0000	7471B	PPB	28 Aug 2020 10:43:34	S	Std
ICV88 - 1	ICV88	19140	3.8996	-	7471B	PPB	28 Aug 2020 10:46:18	U	SMPL
ICB88 - 1	ICB88	-156	-0.0081	-	7471B	PPB	28 Aug 2020 10:48:33	U	SMPL
CCV95 - 1	CCV95	25744	5.2370	-	7471B	PPB	28 Aug 2020 10:50:50	U	SMPL
CCB95 - 1	CCB95	-190	-0.0149	-	7471B	PPB	28 Aug 2020 10:53:05	U	SMPL
CRA - 1	CRA	934	0.2127	-	7471B	PPB	28 Aug 2020 10:55:21	U	SMPL
HighStd - 1	HighStd	49817	10.1121	-	7471B	PPB	28 Aug 2020 10:57:36	U	SMPL
ChkStd - 1	ChkStd	35779	7.2692	-	7471B	PPB	28 Aug 2020 10:59:51	U	SMPL
PB131254BL - 1	PBS	-248	-0.0267	-	7471B	PPB	28 Aug 2020 11:02:09	U	SMPL
PB131254BS - 1	LCSS	17742	3.6165	-	7471B	PPB	28 Aug 2020 11:04:27	U	SMPL
L3507-13 - 1	SU-03-082720	6066	1.2520	-	7471B	PPB	28 Aug 2020 11:06:44	U	SMPL
L3816-01 - 1	ALMASI-1	632	0.1515	-	7471B	PPB	28 Aug 2020 11:09:00	U	SMPL
L3816-01DUP - 1	ALMASI-1DUP	492	0.1232	-	7471B	PPB	28 Aug 2020 11:11:17	U	SMPL
L3816-01MS - 1	ALMASI-1MS	19113	3.8941	-	7471B	PPB	28 Aug 2020 11:13:34	U	SMPL
L3816-01MSD - 1	ALMASI-1MSD	19103	3.8921	-	7471B	PPB	28 Aug 2020 11:15:50	U	SMPL
CCV96 - 1	CCV96	25476	5.1827	-	7471B	PPB	28 Aug 2020 11:18:06	U	SMPL
CCB96 - 1	CCB96	-158	-0.0085	-	7471B	PPB	28 Aug 2020 11:20:22	U	SMPL
L3816-02 - 1	ALMASI-2	598	0.1446	-	7471B	PPB	28 Aug 2020 11:22:39	U	SMPL
L3816-03 - 1	ALMASI-3	400	0.1045	-	7471B	PPB	28 Aug 2020 11:24:54	U	SMPL
L3818-01 - 1	AC-1	3487	0.7297	-	7471B	PPB	28 Aug 2020 11:27:10	U	SMPL
L3818-02 - 1	AC-2	3298	0.6914	-	7471B	PPB	28 Aug 2020 11:29:25	U	SMPL
L3818-03 - 1	AC-3	3014	0.6339	-	7471B	PPB	28 Aug 2020 11:31:42	U	SMPL
L3816-01LX5 - 1		-69	0.0096	-	7471B	PPB	28 Aug 2020 11:33:58	U	SMPL
L3816-01A - 1		18516	3.7732	-	7471B	PPB	28 Aug 2020 11:36:14	U	SMPL
CCV97 - 1	CCV97	24926	5.0713	-	7471B	PPB	28 Aug 2020 11:38:31	U	SMPL
CCB97 - 1	CCB97	-115	0.0002	-	7471B	PPB	28 Aug 2020 11:40:49	U	SMPL