



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110730

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/28/2020	11:00
2	CAL2	1	Water	NA	NA	20.2	6.99	08/28/2020	11:01
3	CAL3	1	Water	NA	NA	20.2	10.03	08/28/2020	11:02
4	ICV	1	Water	NA	NA	20.2	7.00	08/28/2020	11:05
5	CCV1	1	Water	NA	NA	20.2	2.02	08/28/2020	11:06
6	L3814-13	1	Water	NA	NA	15.6	5.05	08/28/2020	11:09
7	L3814-13DUP	1	Water	NA	NA	15.7	5.06	08/28/2020	11:10
8	L3814-14	1	Water	NA	NA	15.6	8.08	08/28/2020	11:11
9	L3814-15	1	Water	NA	NA	15.9	5.48	08/28/2020	11:13
10	L3814-16	1	Water	NA	NA	15.3	5.29	08/28/2020	11:15
11	L3814-17	1	Water	NA	NA	15.6	8.41	08/28/2020	11:16
12	L3814-20	1	Water	NA	NA	14.3	6.58	08/28/2020	11:19
13	L3814-21	1	Water	NA	NA	13.3	6.79	08/28/2020	11:20
14	L3814-22	1	Water	NA	NA	13.3	6.74	08/28/2020	11:22
15	CCV2	1	Water	NA	NA	20.1	12.02	08/28/2020	11:25