



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110733

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/28/2020	11:40
2	L3814-11		1	8	100.00	>212.0	>212.0	08/28/2020	12:05
3	L3814-11DUP		1	8	100.00	>212.0	>212.0	08/28/2020	12:35
4	L3814-19		1	9	100.00	>212.0	>212.0	08/28/2020	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 110733

WorkList Name : fp82820 WorkList ID : 141849 Department : Wet-Chemistry Date : 08-28-2020 09:02:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/10/2020	Water	L3814-11 C	Flash Point	Cool 4 deg C	PSEG01	L51	08-09-12	08/27/2020	1010A
09/10/2020	Water	L3814-19 D	Flash Point	Cool 4 deg C	PSEG01	L51	87-88-89-07-10	08/27/2020	1010A

Date/Time 08/28/2020 11:15
 Received by: KS (welcher)
 Relinquished by: KS (welcher)

Date/Time 08/28/2020 13:15
 Received by: KS (welcher)
 Relinquished by: KS (welcher)