



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : ketankumar

Run Number: LB110737

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/28/2020	13:50
2	CAL2	1	Water	NA	NA	20.2	7.00	08/28/2020	13:51
3	CAL3	1	Water	NA	NA	20.2	10.03	08/28/2020	13:53
4	ICV	1	Water	NA	NA	20.2	6.98	08/28/2020	14:00
5	CCV1	1	Water	NA	NA	21.3	2.02	08/28/2020	14:03
6	L3841-01	1	Water	NA	NA	17.9	5.15	08/28/2020	14:05
7	L3841-01DUP	1	Water	NA	NA	17.8	5.16	08/28/2020	14:06
8	L3841-02	1	Water	NA	NA	18.7	4.97	08/28/2020	14:09
9	L3841-03	1	Water	NA	NA	19.9	5.00	08/28/2020	14:10
10	L3841-04	1	Water	NA	NA	19.3	4.96	08/28/2020	14:11
11	CCV2	1	Water	NA	NA	20.2	12.02	08/28/2020	14:15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 13841 WorkList ID : 141875 Department : Wet-Chemistry Date : 08-28-2020 13:37:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/27/2020	Water	L3841-01-1A	pH	Cool 4 deg C	NYUL01	L51	TH-4-30-DI-1	08/27/2020	SM 4500-H B
08/27/2020	Water	L3841-02-1A	pH	Cool 4 deg C	NYUL01	L51	TH-4-30-DI-2	08/27/2020	SM 4500-H B
08/27/2020	Water	L3841-03-1A	pH	Cool 4 deg C	NYUL01	L51	TH-4-30-DI-3	08/27/2020	SM 4500-H B
08/27/2020	Water	L3841-04-1A	pH	Cool 4 deg C	NYUL01	L51	TH-4-04-DI-4	08/27/2020	SM 4500-H B

Date/Time 08/28/2020 13:40
 Received by: JP (wlf) - Cheny
 Relinquished by: JP

Date/Time 08/28/2020 13:40
 Received by: JP (wlf)
 Relinquished by: JP Cheny