



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110759

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/31/2020	15:35
2	CAL2	1	Water	NA	NA	20.3	7.00	08/31/2020	15:36
3	CAL3	1	Water	NA	NA	20.3	10.03	08/31/2020	15:37
4	ICV	1	Water	NA	NA	20.3	7.00	08/31/2020	15:38
5	CCV1	1	Water	NA	NA	20.3	2.01	08/31/2020	15:39
6	L3854-06	1	Water	NA	NA	23.0	7.83	08/31/2020	15:40
7	L3854-06DUP	1	Water	NA	NA	23.1	7.84	08/31/2020	15:41
8	CCV2	1	Water	NA	NA	20.3	12.02	08/31/2020	15:42

10710759

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3854 W WorkList ID : 141918 Department : Wet-Chemistry Date : 08-31-2020 15:28:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/31/2020	Water	L3854-06	pH	Cool 4 deg C	PSEG01	L62	1332-FM648	08/31/2020	9040C

Date/Time 08.31.2020 16130
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 08.31.2020 15130
 Received by: [Signature]
 Relinquished by: [Signature]