



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110760

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/31/2020	15:35
2	CAL2	1	Water	NA	NA	20.3	7.00	08/31/2020	15:36
3	CAL3	1	Water	NA	NA	20.3	10.03	08/31/2020	15:37
4	ICV	1	Water	NA	NA	20.3	7.00	08/31/2020	15:38
5	CCV1	1	Water	NA	NA	20.3	2.01	08/31/2020	15:39
6	L3854-02	1	Solid	20.03	20	24.0	9.14	08/31/2020	15:45
7	L3854-02DUP	1	Solid	20.04	20	24.1	9.15	08/31/2020	15:46
8	L3854-03	1	Solid	20.02	20	23.8	9.23	08/31/2020	15:49
9	L3854-04	1	Solid	20.01	20	24.0	7.89	08/31/2020	15:50
10	CCV2	1	Water	NA	NA	20.3	12.02	08/31/2020	15:51

WORKLIST(Hardcopy Internal Chain)

LB 110760

WorkList Name: PH L3854 WorkList ID: 141917 Department: Wet-Chemistry Date: 08-31-2020 15:27:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/31/2020	Solid	L3854-02	pH	Cool 4 deg C	PSEG01	L62	DRUM-1	08/31/2020	9045D
08/31/2020	Solid	L3854-03	pH	Cool 4 deg C	PSEG01	L62	DRUM-2	08/31/2020	9045D
08/31/2020	Solid	L3854-04	pH	Cool 4 deg C	PSEG01	L62	DRUM-3	08/31/2020	9045D

Date/Time 08-31-2020 16:30
 Received by: JA wet
 Relinquished by: JA wet

Date/Time 08-31-2020 16:30
 Received by: JA wet
 Relinquished by: JA wet