


LB110762

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.049	0.049	-54	0.000	-54				
0.2	0.200	0.209	0.009	507	0.0 %	507				
2.5	2.500	2.447	-0.053	8325	0.0 %	8325				
5.0	5.000	4.953	-0.047	17078	0.0 %	17078				
7.5	7.500	7.524	0.024	26059	0.0 %	26059				
10.0	10.000	10.017	0.017	34766	0.0 %	34766				

LB110762 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 01 Sep 2020 10:28:03					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-54	-	0.0000	7470A	PPB	01 Sep 2020 10:31:01	S	Std
0.2 - 1	SO.2	507	-	0.2000	7470A	PPB	01 Sep 2020 10:33:19	S	Std
2.5 - 1	SO.5	8325	-	2.5000	7470A	PPB	01 Sep 2020 10:35:36	S	Std
5.0 - 1	SO	17078	-	5.0000	7470A	PPB	01 Sep 2020 10:37:53	S	Std
7.5 - 1	SO.5	26059	-	7.5000	7470A	PPB	01 Sep 2020 10:40:09	S	Std
10.0 - 1	SID	34766	-	10.0000	7470A	PPB	01 Sep 2020 10:42:26	S	Std
ICV92 - 1	ICV92	12950	3.7715	-	7470A	PPB	01 Sep 2020 10:45:19	U	SMPL
ICB92 - 1	ICB92	-259	-0.0101	-	7470A	PPB	01 Sep 2020 10:47:33	U	SMPL
CCV15 - 1	CCV15	17836	5.1703	-	7470A	PPB	01 Sep 2020 10:49:49	U	SMPL
CCB15 - 1	CCB15	-291	-0.0193	-	7470A	PPB	01 Sep 2020 10:52:04	U	SMPL
CRA - 1	CRA	455	0.1943	-	7470A	PPB	01 Sep 2020 10:54:20	U	SMPL
HighStd - 1	HighStd	35628	10.2640	-	7470A	PPB	01 Sep 2020 10:56:36	U	SMPL
ChkStd - 1	ChkStd	24262	7.0100	-	7470A	PPB	01 Sep 2020 10:58:51	U	SMPL
PB131300BL - 1	PBW	-385	-0.0462	-	7470A	PPB	01 Sep 2020 11:01:10	U	SMPL
PB131300BS - 1	LCSW	13805	4.0162	-	7470A	PPB	01 Sep 2020 11:03:28	U	SMPL
L3508-28 - 1	TR-08-082820	-272	-0.0139	-	7470A	PPB	01 Sep 2020 11:05:44	U	SMPL
L3828-01 - 1	WC	-122	0.0291	-	7470A	PPB	01 Sep 2020 11:08:01	U	SMPL
L3837-01 - 1	789UMR-DSP01-01	-103	0.0345	-	7470A	PPB	01 Sep 2020 11:10:18	U	SMPL
L3837-01DUP - 1	789UMR-DSP01-01DUP	-84	0.0400	-	7470A	PPB	01 Sep 2020 11:12:36	U	SMPL
CCV16 - 1	CCV16	17760	5.1485	-	7470A	PPB	01 Sep 2020 11:17:06	U	SMPL
CCB16 - 1	CCB16	-250	-0.0076	-	7470A	PPB	01 Sep 2020 11:19:21	U	SMPL
L3837-02 - 1	L3837-01MS	13091	3.8118	-	7470A	PPB	01 Sep 2020 11:23:53	U	SMPL
L3837-03 - 1	L3837-01MSD	13334	3.8814	-	7470A	PPB	01 Sep 2020 11:26:09	U	SMPL
L3837-04 - 1	789UMR-DSP01-02	-248	-0.0070	-	7470A	PPB	01 Sep 2020 11:28:25	U	SMPL
L3837-05 - 1	789UMR-DSP02-01	-95	0.0368	-	7470A	PPB	01 Sep 2020 11:30:41	U	SMPL
PB131301BL - 1	PBW	-73	0.0431	-	7470A	PPB	01 Sep 2020 11:32:58	U	SMPL
PB131301BS - 1	LCSW	12547	3.6561	-	7470A	PPB	01 Sep 2020 11:35:14	U	SMPL
L3828-05 - 1	GW	21730	6.2851	-	7470A	PPB	01 Sep 2020 11:37:31	U	SMPL
L3829-01 - 1	F06045	42	0.0760	-	7470A	PPB	01 Sep 2020 11:51:52	U	SMPL
L3829-01DUP - 1	F06045DUP	39	0.0752	-	7470A	PPB	01 Sep 2020 11:54:07	U	SMPL
CCV17 - 1	CCV17	17672	5.1233	-	7470A	PPB	01 Sep 2020 11:56:23	U	SMPL
CCB17 - 1	CCB17	-163	0.0174	-	7470A	PPB	01 Sep 2020 11:58:38	U	SMPL
L3829-01MS - 1	F06045MS	14651	4.2585	-	7470A	PPB	01 Sep 2020 12:00:54	U	SMPL
L3829-01MSD - 1	F06045MSD	15190	4.4128	-	7470A	PPB	01 Sep 2020 12:03:10	U	SMPL
L3831-19 - 1	RB-01-20200827	-129	0.0271	-	7470A	PPB	01 Sep 2020 12:05:25	U	SMPL
PB131300TB - 1	PB131280TB	-71	0.0437	-	7470A	PPB	01 Sep 2020 12:07:42	U	SMPL
L3837-01LX5 - 1		-34	0.0543	-	7470A	PPB	01 Sep 2020 12:09:58	U	SMPL
L3837-01A - 1		13500	3.9289	-	7470A	PPB	01 Sep 2020 12:12:14	U	SMPL
L3829-01LX5 - 1		-127	0.0277	-	7470A	PPB	01 Sep 2020 12:14:31	U	SMPL
L3829-01A - 1		13725	3.9933	-	7470A	PPB	01 Sep 2020 12:16:48	U	SMPL
CCV18 - 1	CCV18	17951	5.2032	-	7470A	PPB	01 Sep 2020 12:19:05	U	SMPL
CCB18 - 1	CCB18	-201	0.0065	-	7470A	PPB	01 Sep 2020 12:21:21	U	SMPL