



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110775

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/01/2020	15:50
2	L3854-07		1	8	100.00	>212.0	>212.0	09/01/2020	16:25
3	L3854-07DUP		1	8	100.00	>212.0	>212.0	09/01/2020	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB110775

WorkList Name : FP9120 WorkList ID : 141958 Department : Wet-Chemistry Date : 09-01-2020 14:03:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/03/2020	Water	L3854-07	C Flash Point	Cool 4 deg C	PSEG01	L62	1332-FM648	08/31/2020	1010A

Date/Time 09/01/2020 15:25
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)

Date/Time 09/01/2020 16:45
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)