



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110777

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/01/2020	17:20
2	CAL2	1	Water	NA	NA	20.3	7.01	09/01/2020	17:21
3	CAL3	1	Water	NA	NA	20.3	10.03	09/01/2020	17:22
4	ICV	1	Water	NA	NA	20.3	7.00	09/01/2020	17:25
5	CCV1	1	Water	NA	NA	20.3	2.01	09/01/2020	17:26
6	L3866-05	1	Solid	20.03	20	25.4	8.66	09/01/2020	17:29
7	L3866-05DUP	1	Solid	20.04	20	25.5	8.67	09/01/2020	17:30
8	L3873-01	1	Solid	20.02	20	25.3	8.42	09/01/2020	17:31
9	CCV2	1	Water	NA	NA	20.2	12.02	09/01/2020	17:33

VB 110777

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L3878 WorkList ID : 141961 Department : Wet-Chemistry Date : 09-01-2020 14:16:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/01/2020	Solid	L3866-05	Corrosivity	Cool 4 deg C	PSEG05	M11	OK-03-090120	09/01/2020	9045D
09/01/2020	Solid	L3873-01	Corrosivity	Cool 4 deg C	PSEG01	M11	DRUM-1-2	09/01/2020	9045D

Date/Time 09.01.2020 16:10
 Received by: JA (wet chemistry)
 Relinquished by: KS (wet chemistry)

Date/Time 09.01.2020 17:50
 Received by: KS (wet chemistry)
 Relinquished by: JA (wet chemistry)
 Chen