



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110780

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	3.99	09/01/2020	17:25
2	CAL2	1	Water	NA	NA	21.1	7.00	09/01/2020	17:26
3	CAL3	1	Water	NA	NA	21.1	10.03	09/01/2020	17:27
4	ICV	1	Water	NA	NA	21.2	7.01	09/01/2020	17:29
5	CCV1	1	Water	NA	NA	21.2	2.01	09/01/2020	17:30
6	L3858-02	1	Water	NA	NA	19.1	6.90	09/01/2020	17:32
7	L3858-02DUP	1	Water	NA	NA	19.2	6.91	09/01/2020	17:33
8	L3874-01	1	Water	NA	NA	26.0	5.96	09/01/2020	17:34
9	L3876-14	1	Water	NA	NA	25.3	6.69	09/01/2020	17:35
10	CCV2	1	Water	NA	NA	21.1	12.02	09/01/2020	17:36

WORKLIST(Hardcopy Internal Chain)

W3110780

WorkList Name : PH L3874 WorkList ID : 141964 Department : Wet-Chemistry Date : 09-01-2020 15:30:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/01/2020	Water	L3858-02	pH	Cool 4 deg C	PARS15	L41	HAM-WC-L-VB01	09/01/2020	9040C
09/01/2020	Water	L3874-01	pH	Cool 4 deg C	PSEG01	M12	31216	09/01/2020	9040C
09/01/2020	Water	L3876-14	pH	Cool 4 deg C	PSEG01	M11	1576	09/01/2020	9040C

Date/Time 09.01.2020 16:10
 Received by: Jo [signature]
 Relinquished by: Jo [signature]

Date/Time 09.01.2020 17:35
 Received by: Jo [signature]
 Relinquished by: Jo [signature]