



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110785

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/02/2020	09:35
2	L3858-02		1	8	100.00	>212.0	>212.0	09/02/2020	10:05
3	L3858-02DUP		1	8	100.00	>212.0	>212.0	09/02/2020	10:40
4	L3874-01		1	9	100.00	>212.0	>212.0	09/02/2020	11:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp9220

WorkList ID : 141994

Department : Wet-Chemistry

Date : 09-02-2020 08:59:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/03/2020	Water	L3858-02	Flash Point	Cool 4 deg C	PARS15	L41	HAM-WC-L-VB01	09/01/2020	1010A
09/08/2020	Water	L3874-01	Flash Point	Cool 4 deg C	PSEG01	M12	31216	09/01/2020	1010A

Date/Time 09/02/2020 09:02
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 09/02/2020 11:25
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)