



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110793

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.00	09/02/2020	15:00
2	CAL2	1	Water	NA	NA	20.1	7.01	09/02/2020	15:01
3	CAL3	1	Water	NA	NA	20.1	10.03	09/02/2020	15:02
4	ICV	1	Water	NA	NA	20.2	7.00	09/02/2020	15:05
5	CCV1	1	Water	NA	NA	20.1	2.01	09/02/2020	15:06
6	L3886-03	1	Water	NA	NA	25.6	7.06	09/02/2020	15:25
7	L3886-03DUP	1	Water	NA	NA	25.7	7.07	09/02/2020	15:26
8	CCV2	1	Water	NA	NA	20.1	12.03	09/02/2020	15:30

LB 110793

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3887 W WorkList ID : 142014 Department : Wet-Chemistry Date : 09-02-2020 14:43:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/02/2020	Water	L3886-03	pH	Cool 4 deg C	PSEG01	M13	TOTE-31231	09/02/2020	9040C

Date/Time 09/02/2020 14:30
 Received by: JA (wet chem)
 Relinquished by: KS (wet chem)

Date/Time 09/02/2020 17:45
 Received by: KS (wet chem)
 Relinquished by: JA (wet chem)