


LB110794

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.2650e-004 *Slope*C= -6.0473e-002 *y-intercept*

Rho= 0.9999135

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.033	0.033	413	0.000	413					0
0.05	0.050					159					1
0.20	0.200	0.237	0.037	1313	0.0 %	1313					19
2.50	2.500	2.454	-0.046	11102	0.0 %	11102					-2
5.00	5.000	4.961	-0.039	22170	0.0 %	22170					-1
7.50	7.500	7.444	-0.056	33135	0.0 %	33135					-1
10.0	10.000	10.070	0.070	44729	0.0 %	44729					1

LB110794
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	413	-		0 ISM02.4	PPB	9/2/2020 15:20	Std
	0.2 S0.2	1313	-		0.2 ISM02.4	PPB	9/2/2020 15:23	Std
	2.5 S2.5	11102	-		2.5 ISM02.4	PPB	9/2/2020 15:25	Std
	5 S5.0	22170	-		5 ISM02.4	PPB	9/2/2020 15:27	Std
	7.5 S7.5	33135	-		7.5 ISM02.4	PPB	9/2/2020 15:29	Std
	10 S10.0	44729	-		10 ISM02.4	PPB	9/2/2020 15:32	Std
ICV	ICV	17727	3.9546 -		ISM02.4	PPB	9/2/2020 15:38	SMPL
ICB	ICB	350	0.0188 -		ISM02.4	PPB	9/2/2020 15:40	SMPL
CCV038	CCV038	22768	5.0964 -		ISM02.4	PPB	9/2/2020 15:42	SMPL
CCB038	CCB038	300	0.0075 -		ISM02.4	PPB	9/2/2020 15:45	SMPL
PB131379BL	PBS379	344	0.0174 -		ISM02.4	PPB	9/2/2020 15:47	SMPL
L3844-01	MCOAA3	5615	1.2113 -		ISM02.4	PPB	9/2/2020 15:49	SMPL
L3844-02	MCOAA4	4539	0.9676 -		ISM02.4	PPB	9/2/2020 15:51	SMPL
L3844-03	MCOAA5	7580	1.6564 -		ISM02.4	PPB	9/2/2020 15:54	SMPL
L3844-04	MCOAA6	6750	1.4684 -		ISM02.4	PPB	9/2/2020 15:56	SMPL
L3844-05	MCOAA7	6194	1.3424 -		ISM02.4	PPB	9/2/2020 15:58	SMPL
L3844-06	MCOAD2	1062	0.1801 -		ISM02.4	PPB	9/2/2020 16:00	SMPL
L3844-07	MCOAE0	5277	1.1347 -		ISM02.4	PPB	9/2/2020 16:03	SMPL
L3844-08	MCOAE3	1004	0.1669 -		ISM02.4	PPB	9/2/2020 16:05	SMPL
L3844-09	MCOAE5	1050	0.1773 -		ISM02.4	PPB	9/2/2020 16:07	SMPL
L3844-12	MCOAA8	3905	0.824 -		ISM02.4	PPB	9/2/2020 16:10	SMPL
L3844-13	MCOAA9	6162	1.3352 -		ISM02.4	PPB	9/2/2020 16:12	SMPL
L3844-14	MCOAB0	1894	0.3685 -		ISM02.4	PPB	9/2/2020 16:14	SMPL
L3844-15	MCOAB1	1933	0.3773 -		ISM02.4	PPB	9/2/2020 16:16	SMPL
L3844-16	MCOAB2	3792	0.7984 -		ISM02.4	PPB	9/2/2020 16:19	SMPL
L3844-17	MCOAB3	4661	0.9952 -		ISM02.4	PPB	9/2/2020 16:21	SMPL
L3844-18	MCOAB5	4840	1.0358 -		ISM02.4	PPB	9/2/2020 16:23	SMPL
L3844-19	MCOAB7	7564	1.6527 -		ISM02.4	PPB	9/2/2020 16:25	SMPL
CCV039	CCV039	22180	4.9632 -		ISM02.4	PPB	9/2/2020 16:28	SMPL
CCB039	CCB039	338	0.0161 -		ISM02.4	PPB	9/2/2020 16:30	SMPL
L3844-20	MCOAB8	6435	1.397 -		ISM02.4	PPB	9/2/2020 16:32	SMPL
L3844-21	MCOAB8D	6273	1.3603 -		ISM02.4	PPB	9/2/2020 16:34	SMPL
L3844-22	MCOAB8S	17871	3.9872 -		ISM02.4	PPB	9/2/2020 16:39	SMPL
CCV040	CCV040	23033	5.1564 -		ISM02.4	PPB	9/2/2020 16:41	SMPL
CCB040	CCB040	315	0.0109 -		ISM02.4	PPB	9/2/2020 16:44	SMPL