



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 09/04/2020OVEN IN TEMP1: 98 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 09/03/2020 16:00 DATE & TIME OUT1: 09/04/2020 07:15OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 09/04/2020 08:00 DATE & TIME OUT2: 09/04/2020 09:35Run Number: LB110806BalanceID: WC SC-3OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB110806BL	LB110806BL	148.6537	148.6537	100	148.6537	148.6537	148.6537	0.0000	0
2	L3887-03	CONCRETE-PILE	152.7635	152.7635	100	152.8074	152.8074	152.8074	0.0439	439
3	L3887-03DUP	CONCRETE-PILEDUP	143.6079	143.6079	100	143.6514	143.6514	143.6514	0.0435	435

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

VB 210806

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM TS L3887 WorkList ID : 142036 Department : Wet-Chemistry Date : 09-03-2020 11:35:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/07/2020	Solid	L3887-03	ASTM TS	Cool 4 deg C	PSEG01	M13	CONCRETE-PILE	09/02/2020	SM2540 B

Date/Time 09/03/2020 15:40
 Received by: Jo [Signature]
 Relinquished by: Jo [Signature]

Date/Time 09/03/2020 16:20
 Received by: Jo [Signature]
 Relinquished by: Jo [Signature]