


LB110809

245.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 3.1316e-004

C= 1.8563e-002

Rho= 0.9999243

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	-77	0.000	-77				
0.05	0.050	0.064	0.014	146	0.0 %	146				
0.2	0.200	0.202	0.002	586	0.0 %	586				
2.5	2.500	2.526	0.026	8006	0.0 %	8006				
5.0	5.000	4.908	-0.092	15612	0.0 %	15612				
7.5	7.500	7.570	0.070	24115	0.0 %	24115				
10.0	10.000	9.985	-0.015	31826	0.0 %	31826				

LB110809 INSTRUMENT ID: CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Sep 2020 13:47:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-77	-	0.0000	245.1	PPB	03 Sep 2020 13:49:19	S	Std
0.05 - 1	50.05	146	-	0.0500	245.1	PPB	03 Sep 2020 13:51:36	S	Std
0.2 - 1	50.2	586	-	0.2000	245.1	PPB	03 Sep 2020 13:53:53	S	Std
2.5 - 1	52.5	8006	-	2.5000	245.1	PPB	03 Sep 2020 13:56:09	S	Std
5.0 - 1	55	15612	-	5.0000	245.1	PPB	03 Sep 2020 13:58:26	S	Std
7.5 - 1	57.5	24115	-	7.5000	245.1	PPB	03 Sep 2020 14:00:42	S	Std
10.0 - 1	510	31826	-	10.0000	245.1	PPB	03 Sep 2020 14:02:59	S	Std
ICV97 - 1	ICV97	12556	3.9506	-245.1		PPB	03 Sep 2020 14:09:10	U	SMPL
ICB97 - 1	ICB97	-74	-0.0046	-245.1		PPB	03 Sep 2020 14:14:09	U	SMPL
CCV46 - 1	CCV46	16013	5.0332	-245.1		PPB	03 Sep 2020 14:16:25	U	SMPL
CCB46 - 1	CCB46	-75	-0.0049	-245.1		PPB	03 Sep 2020 14:21:24	U	SMPL
CRA - 1	CRA	118	0.0555	-245.1		PPB	03 Sep 2020 14:23:38	U	SMPL
HighStd - 1	HighStd	31966	10.0291	-245.1		PPB	03 Sep 2020 14:25:53	U	SMPL
ChkStd - 1	ChkStd	22060	6.9269	-245.1		PPB	03 Sep 2020 14:28:09	U	SMPL
PB131399BL - 1	PBW	-65	-0.0018	-245.1		PPB	03 Sep 2020 14:33:34	U	SMPL
PB131399BS - 1	LCSW	11483	3.6146	-245.1		PPB	03 Sep 2020 14:35:49	U	SMPL
L3893-02 - 1	COMP	580	0.2002	-245.1		PPB	03 Sep 2020 14:38:05	U	SMPL
L3893-02DUP - 1	COMPDUP	533	0.1855	-245.1		PPB	03 Sep 2020 14:42:54	U	SMPL
L3893-02MS - 1	COMPMS	11153	3.5113	-245.1		PPB	03 Sep 2020 14:47:43	U	SMPL
L3893-02MSD - 1	COMPMSD	11489	3.6165	-245.1		PPB	03 Sep 2020 14:50:00	U	SMPL
CCV47 - 1	CCV47	15783	4.9612	-245.1		PPB	03 Sep 2020 14:52:16	U	SMPL
CCB47 - 1	CCB47	-214	-0.0485	-245.1		PPB	03 Sep 2020 14:54:31	U	SMPL
L3893-02LX5 - 1		64	0.0386	-245.1		PPB	03 Sep 2020 14:56:48	U	SMPL
L3893-02A - 1		11694	3.6807	-245.1		PPB	03 Sep 2020 14:59:03	U	SMPL
CCV48 - 1	CCV48	15676	4.9277	-245.1		PPB	03 Sep 2020 15:01:18	U	SMPL
CCB48 - 1	CCB48	-215	-0.0488	-245.1		PPB	03 Sep 2020 15:03:35	U	SMPL