


LB110825

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3479e-004

C= -7.0666e-002

Rho= 0.9999679

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.015	0.015	365	0.000	365					11
0.05	0.050					159					12
0.20	0.200	0.223	0.023	1250	0.0 %	1250					11
2.50	2.500	2.485	-0.015	10884	0.0 %	10884					11
5.00	5.000	4.940	-0.060	21342	0.0 %	21342					11
7.50	7.500	7.518	0.018	32320	0.0 %	32320					0
10.0	10.000	10.019	0.019	42975	0.0 %	42975					0

LB110825
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	365	-		0 ISM02.4	PPB	9/4/2020 11:14	Std
	0.2 S0.2	1250	-		0.2 ISM02.4	PPB	9/4/2020 11:16	Std
	2.5 S2.5	10884	-		2.5 ISM02.4	PPB	9/4/2020 11:19	Std
	5 S5.0	21342	-		5 ISM02.4	PPB	9/4/2020 11:21	Std
	7.5 S7.5	32320	-		7.5 ISM02.4	PPB	9/4/2020 11:23	Std
	10 S10.0	42975	-		10 ISM02.4	PPB	9/4/2020 11:26	Std
ICV	ICV	17184	3.964 -		ISM02.4	PPB	9/4/2020 11:28	SMPL
ICB	ICB	260	-0.0096 -		ISM02.4	PPB	9/4/2020 11:31	SMPL
CCV058	CCV058	21975	5.0888 -		ISM02.4	PPB	9/4/2020 11:33	SMPL
CCB058	CCB058	240	-0.0143 -		ISM02.4	PPB	9/4/2020 11:35	SMPL
PB131421BL	PBS421	-18	-0.0749 -		ISM02.4	PPB	9/4/2020 11:37	SMPL
L3845-01	MC0AC0	168	-0.0312 -		ISM02.4	PPB	9/4/2020 11:42	SMPL
L3845-02	MC0AC1	2546	0.5271 -		ISM02.4	PPB	9/4/2020 11:44	SMPL
L3845-03	MC0AD3	8085	1.8276 -		ISM02.4	PPB	9/4/2020 11:47	SMPL
L3845-04	MC0AE1	4626	1.0155 -		ISM02.4	PPB	9/4/2020 11:49	SMPL
L3845-05	MC0AD0	7302	1.6438 -		ISM02.4	PPB	9/4/2020 11:51	SMPL
L3845-06	MC0AC8	10238	2.3331 -		ISM02.4	PPB	9/4/2020 11:54	SMPL
L3845-07	MC0AC6	25223	5.8514 -		ISM02.4	PPB	9/4/2020 11:56	SMPL
L3845-10	MC0AC6D	26009	6.036 -		ISM02.4	PPB	9/4/2020 11:58	SMPL
L3845-11	MC0AC6S	34677	8.0711 -		ISM02.4	PPB	9/4/2020 12:00	SMPL
CCV059	CCV059	22471	5.2053 -		ISM02.4	PPB	9/4/2020 12:03	SMPL
CCB059	CCB059	260	-0.0096 -		ISM02.4	PPB	9/4/2020 12:05	SMPL