



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 09/04/2020

Run Number: LB110826

BalanceID: WC SC-3

OvenID: WC OVEN-1

FilterID: 17023242

ThermometerID: Wet Oven#1

TEMP1 IN: 104 °C 09/04/2020 07:15 TEMP1 OUT: 104 °C 09/04/2020 08:20
TEMP2 IN: 103 °C 09/04/2020 08:45 TEMP2 OUT: 103 °C 09/04/2020 10:00
TEMP3 IN: 103 °C 09/04/2020 11:20 TEMP3 OUT: 103 °C 09/04/2020 12:45
TEMP4 IN: 103 °C 09/04/2020 13:25 TEMP4 OUT: 103 °C 09/04/2020 15:35

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB110826BL	LB110826BL	2.4563	2.4563	100	2.4563	2.4563	2.4563	0.0000	0
2	LB110826BS	LB110826BS	1.3205	1.3205	100	1.3752	1.3752	1.3752	0.0547	547
3	L3883-01	001-WILLETTS-PT-BLVD- (SEP)	1.3737	1.3737	150	1.3782	1.3782	1.3782	0.0045	30
4	L3883-02	002-35TH-AVE- (SEP)	1.3648	1.3648	150	1.3702	1.3702	1.3702	0.0054	36
5	L3909-01	32220008-3	1.3786	1.3786	500	1.3831	1.3831	1.3831	0.0045	9
6	L3910-01	LPBH-01	1.3733	1.3733	600	1.3779	1.3779	1.3779	0.0046	7.7
7	L3910-01DUP	LPBH-01DUP	1.3722	1.3722	600	1.3769	1.3769	1.3769	0.0047	7.8
8	L3921-01	UDS-4-26	1.3815	1.3815	1000	1.4061	1.4061	1.4061	0.0246	24.6
9	L3922-01	UDS-6-16A	1.3805	1.3805	800	1.3935	1.3935	1.3935	0.0130	16.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

110826

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss L3910 WorkList ID : 142054 Department : Wet-Chemistry Date : 09-04-2020 07:56:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/07/2020	Water	L3883-01	TSS	Cool 4 deg C	TULL01	M11	001-WILLETTS-PT-BLVD-(SEP)	09/01/2020	SM2540 D
09/07/2020	Water	L3883-02	TSS	Cool 4 deg C	TULL01	M11	002-35TH-AVE-(SEP)	09/01/2020	SM2540 D
09/10/2020	Water	L3909-01	TSS	Cool 4 deg C	PHAR02	M31	32220008-3	09/03/2020	SM2540 D
09/07/2020	Water	L3910-01	TSS	Cool 4 deg C	DVIR01	M51	LPBH-01	09/02/2020	SM2540 D
09/10/2020	Water	L3921-01	TSS	Cool 4 deg C	LIRO01	M31	UDS-4-26	09/03/2020	SM2540 D
09/08/2020	Water	L3922-01	TSS	Cool 4 deg C	LIRO01	M51	UDS-6-16A	09/03/2020	SM2540 D

Date/Time 09/10/2020 10:10
Received by: Jo [Signature]
Relinquished by: Jo [Signature]

Date/Time 09/04/2020 11:35
Received by: Jo [Signature]
Relinquished by: Jo [Signature]