



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110827

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	09/04/2020	11:15
2	CAL2	1	Water	NA	NA	21.3	7.00	09/04/2020	11:16
3	CAL3	1	Water	NA	NA	21.3	10.03	09/04/2020	11:19
4	ICV	1	Water	NA	NA	21.4	7.00	09/04/2020	11:20
5	CCV1	1	Water	NA	NA	21.3	2.01	09/04/2020	11:22
6	L3920-01	1	Water	NA	NA	25.5	6.95	09/04/2020	11:25
7	L3920-01DUP	1	Water	NA	NA	25.6	6.96	09/04/2020	11:26
8	CCV2	1	Water	NA	NA	21.3	12.02	09/04/2020	11:27

HP 110827

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3920 WorkList ID : 142088 Department : Wet-Chemistry Date : 09-04-2020 11:06:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/03/2020	Water	L3920-01	pH	Cool 4 deg C	STER08	M31	POOL-WATER	09/03/2020	SM 4500-H B

Date/Time 09/04/2020 11:10
Received by: Jo / Jost
Relinquished by: Jo / Jost

Date/Time 09-04-2020 12:12
Received by: Jo / Jost
Relinquished by: Jo / Jost