



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110828

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP86658
Turbidity Calibration std, 20NTU	WP86654
Turbidity Calibration std, 40NTU	WP86653
Turbidity Calibration std, 1NTU	WP86657
Turbidity Calibration std, 5NTU	WP86656
Turbidity Calibration - CCV std, 10 NTU	WP86655

Intercept: 0.1576

Slope: 0.9658

Regression: 0.9994

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.235	09/04/2020	11:01
2	CAL2	1	1	1.180	09/04/2020	11:03
3	CAL3	5	1	5.250	09/04/2020	11:05
4	CAL4	10	1	9.980	09/04/2020	11:07
5	CAL5	20	1	18.500	09/04/2020	11:09
6	CAL6	40	1	39.200	09/04/2020	11:12



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB110828

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.400	10.605	09/04/2020	11:15
2	ICB		1	0.248	0.094	09/04/2020	11:17
3	CCV1	10	1	9.920	10.108	09/04/2020	11:19
4	CCB1		1	0.252	0.098	09/04/2020	11:21
5	LB110828BL		1	0.253	0.099	09/04/2020	11:23
6	L3922-01		1	14.600	14.954	09/04/2020	11:26
7	L3922-01DUP		1	15.700	16.093	09/04/2020	11:29
8	CCV2	10	1	9.930	10.118	09/04/2020	11:32
9	CCB2		1	0.265	0.111	09/04/2020	11:35

18110828 WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-90420 WorkList ID : 142080 Department : Wet-Chemistry Date : 09-04-2020 10:09:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2020	Water	L3922-01	Turbidity	Cool 4 deg C	LIRO01	M51	UDS-6-16A	09/03/2020	SM2130 B

Date/Time 9/04/2020 10:30
 Received by: Am (wet-chem)
 Relinquished by: Am (wet-chem)

Date/Time 9/04/2020 11:50
 Received by: Am (wet-chem)
 Relinquished by: Am (wet-chem)