



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110837

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/04/2020	16:25
2	CAL2	1	Water	NA	NA	20.1	7.00	09/04/2020	16:26
3	CAL3	1	Water	NA	NA	20.1	10.03	09/04/2020	16:27
4	ICV	1	Water	NA	NA	20.2	7.02	09/04/2020	16:30
5	CCV1	1	Water	NA	NA	20.3	2.01	09/04/2020	16:31
6	L3917-01	1	Water	NA	NA	26.2	6.88	09/04/2020	16:45
7	L3917-01DUP	1	Water	NA	NA	26.3	6.89	09/04/2020	16:46
8	CCV2	1	Water	NA	NA	20.3	12.02	09/04/2020	16:48

2B 110837

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3917 WorkList ID : 142103 Department : Wet-Chemistry Date : 09-04-2020 15:23:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/04/2020	Water	L3917-01	pH	Cool 4 deg C	PSEG01	M31	BALER-BLVD-DRUM	09/04/2020	9040C

Date/Time 09-04-2020 15:25
Received by: JG (Wet-Chem)
Relinquished by: RC (Ext Lab)

Date/Time 09-04-2020/6:30
Received by: RC (Ext Lab)
Relinquished by: JG (Wet-Chem)