



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110839

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/04/2020	16:25
2	CAL2	1	Water	NA	NA	20.1	7.00	09/04/2020	16:26
3	CAL3	1	Water	NA	NA	20.1	10.03	09/04/2020	16:27
4	ICV	1	Water	NA	NA	20.2	7.02	09/04/2020	16:30
5	CCV1	1	Water	NA	NA	20.3	2.01	09/04/2020	16:31
6	L3915-01	1	Solid	20.03	20	25.6	8.19	09/04/2020	16:32
7	L3915-01DUP	1	Solid	20.04	20	25.7	8.19	09/04/2020	16:33
8	L3915-05	1	Solid	20.02	20	25.7	8.26	09/04/2020	16:35
9	L3915-09	1	Solid	20.03	20	25.8	8.91	09/04/2020	16:36
10	L3915-13	1	Solid	20.04	20	25.5	8.65	09/04/2020	16:39
11	CCV2	1	Water	NA	NA	20.3	12.02	09/04/2020	16:40

JB 11039

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3915 WorkList ID : 142097 Department : Wet-Chemistry Date : 09-04-2020 13:06:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/03/2020	Solid	L3915-01	pH	Cool 4 deg C	ENTA05	M51	WC-A1-0-4-C	09/03/2020	9045D
09/03/2020	Solid	L3915-05	pH	Cool 4 deg C	ENTA05	M51	WC-PME-0-5-C	09/03/2020	9045D
09/03/2020	Solid	L3915-09	pH	Cool 4 deg C	ENTA05	M51	WC-B2-0-3.5-C	09/03/2020	9045D
09/03/2020	Solid	L3915-13	pH	Cool 4 deg C	ENTA05	M51	WC-A2-10-14-C	09/03/2020	9045D

Date/Time 09.04.2020 15:35
Received by: Jo [Signature]
Relinquished by: Jo [Signature]

Date/Time 09.04.2020 17:10
Received by: Jo [Signature]
Relinquished by: Jo [Signature]