



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110859

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/08/2020	10:58
2	L3886-01		1	8	100.00	>212.0	>212.0	09/08/2020	11:32
3	L3886-01DUP		1	8	100.00	>212.0	>212.0	09/08/2020	12:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 110859

WorkList Name : fp9320 WorkList ID : 142031 Department : Wet-Chemistry Date : 09-03-2020 09:36:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/07/2020	Solid	L3886-01	Flash Point	Cool 4 deg C	PSEG01	M13	TOTE-31230	09/02/2020	1010A

Date/Time 09/08/2020 10:30
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 09/08/2020 12:10
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)