



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB110876

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP86527
COD calibration std. 100 ppm	WP86524
COD calibration std. 10 ppm	WP86522
COD calibration std. 50 ppm	WP86523
COD calibration std. 0 ppm	WP86521
COD calibration std. 150 ppm	WP86525
COD Digestion Vials Low Level 0-150Mg/L	W2735
COD ICV-LCS std, 50ppm	WP86619
COD CCV std, 50ppm	WP86618

Temp In (C): <u>150</u>	Date In: <u>09/09/2020</u>	Time In: <u>09:50</u>
Temp Out (C): <u>150</u>	Date Out: <u>09/09/2020</u>	Time Out: <u>11:50</u>

Intercept: 0.1159

Slope: 1.0078

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	08/27/2020	16:40
2	CAL2	10	1	Water	11.000	08/27/2020	16:40
3	CAL3	50	1	Water	49.000	08/27/2020	16:41
4	CAL4	100	1	Water	102.000	08/27/2020	16:41
5	CAL5	150	1	Water	151.000	08/27/2020	16:42



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB110876

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	52.000	51.483	08/27/2020	16:44
2	ICB				1	Water	0.000	-0.115	08/27/2020	16:44
3	CCV1	50			1	Water	49.000	48.506	09/09/2020	14:40
4	CCB1				1	Water	0.000	-0.115	09/09/2020	14:41
5	LB110876BL				1	Water	0.000	-0.115	09/09/2020	14:41
6	LB110876BS	50			1	Water	52.000	51.483	09/09/2020	14:42
7	L3915-03				1	Water	16.000	15.761	09/09/2020	14:43
8	L3915-03DUP				1	Water	19.000	18.738	09/09/2020	14:43
9	L3915-03MS	50			1	Water	69.000	68.351	09/09/2020	14:44
10	L3915-03MSD	50			1	Water	66.000	65.374	09/09/2020	14:44
11	L3915-07				1	Water	40.000	39.575	09/09/2020	14:45
12	L3915-11				1	Water	42.000	41.560	09/09/2020	14:45
13	L3915-15				1	Water	20.000	19.730	09/09/2020	14:46
14	CCV2	50			1	Water	49.000	48.506	09/09/2020	14:47
15	CCB2				1	Water	0.000	-0.115	09/09/2020	14:47