



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110884

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	09/09/2020	15:00
2	CAL2	1	Water	NA	NA	20.4	7.01	09/09/2020	15:01
3	CAL3	1	Water	NA	NA	20.4	10.03	09/09/2020	15:02
4	ICV	1	Water	NA	NA	20.4	7.00	09/09/2020	15:05
5	CCV1	1	Water	NA	NA	20.5	2.01	09/09/2020	15:06
6	L3941-04	1	Water	NA	NA	26.7	8.62	09/09/2020	15:10
7	L3942-02	1	Water	NA	NA	22.6	7.63	09/09/2020	15:11
8	L3942-02DUP	1	Water	NA	NA	22.7	7.64	09/09/2020	15:12
9	CCV2	1	Water	NA	NA	20.5	12.02	09/09/2020	15:15

2B110884

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L3941 WorkList ID : 142173 Department : Wet-Chemistry Date : 09-09-2020 14:26:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/09/2020	Water	L3941-04-A	Corrosivity	Cool 4 deg C	PSEG01	M41	RBR-200031	09/09/2020	9040C
09/09/2020	Water	L3942-02	Corrosivity	Cool 4 deg C	PSEG01	M41	N35868	09/09/2020	9040C

Date/Time 09-09-2020 14:30
 Received by: JD
 Relinquished by: JD chem

Date/Time 09-09-2020 15:40
 Received by: JD
 Relinquished by: JD chem