



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110895

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/10/2020	12:30
2	CAL2	1	Water	NA	NA	20.3	7.00	09/10/2020	12:31
3	CAL3	1	Water	NA	NA	20.3	10.03	09/10/2020	12:32
4	ICV	1	Water	NA	NA	20.4	6.98	09/10/2020	12:33
5	CCV1	1	Water	NA	NA	20.3	2.03	09/10/2020	12:35
6	L3959-01	1	Water	NA	NA	25.0	5.64	09/10/2020	12:36
7	L3959-01DUP	1	Water	NA	NA	25.1	5.65	09/10/2020	12:37
8	L3960-01	1	Water	NA	NA	24.8	5.05	09/10/2020	12:40
9	L3960-02	1	Water	NA	NA	24.8	4.89	09/10/2020	12:42
10	L3960-03	1	Water	NA	NA	24.6	5.85	09/10/2020	12:45
11	CCV2	1	Water	NA	NA	20.4	12.02	09/10/2020	12:47

VB 110895

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 13960 WorkList ID : 142204 Department : Wet-Chemistry Date : 09-10-2020 12:24:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/09/2020	Water	L3959-01	pH	Cool 4 deg C	STER08	L12	POOL-WATER	09/09/2020	SM 4500-H B
09/09/2020	Water	L3960-01	pH	Cool 4 deg C	STER07	L12	CELLS-A	09/09/2020	SM 4500-H B
09/09/2020	Water	L3960-02	pH	Cool 4 deg C	STER07	L12	CELLS-B/C	09/09/2020	SM 4500-H B
09/09/2020	Water	L3960-03	pH	Cool 4 deg C	STER07	L12	CELLS-D	09/09/2020	SM 4500-H B

Date/Time 09-10-2020 12:24:36
 Received by: ID (wet-chem)
 Relinquished by: AMLwet-chem

Date/Time 09-10-2020 13:00
 Received by: AMLwet-chem
 Relinquished by: ID (wet-chem)
 Chem