



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110897

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/10/2020	12:30
2	CAL2	1	Water	NA	NA	20.3	7.00	09/10/2020	12:31
3	CAL3	1	Water	NA	NA	20.3	10.03	09/10/2020	12:32
4	ICV	1	Water	NA	NA	20.4	6.98	09/10/2020	12:33
5	CCV1	1	Water	NA	NA	20.3	2.03	09/10/2020	12:35
6	L3970-01	1	Solid	20.03	20	25.7	6.58	09/10/2020	13:05
7	L3970-01DUP	1	Solid	20.04	20	25.8	6.59	09/10/2020	13:06
8	L3970-02	1	Solid	20.02	20	26.1	6.87	09/10/2020	13:10
9	CCV2	1	Water	NA	NA	20.3	12.02	09/10/2020	13:11

WORKLIST(Hardcopy Internal Chain)

110897

WorkList Name : PH L3970 WorkList ID : 142205 Department : Wet-Chemistry Date : 09-10-2020 12:43:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/10/2020	Solid	L3970-01	pH	Cool 4 deg C	HOME01	P12	FRONT-SOIL	09/10/2020	9045D
09/10/2020	Solid	L3970-02	pH	Cool 4 deg C	HOME01	P12	BACK-SOIL	09/10/2020	9045D

Date/Time 09-10-2020 13:00
 Received by: JN (wft them)
 Relinquished by: CP (SM 1)

Date/Time 09-10-2020 14:00
 Received by: CP (SM 1)
 Relinquished by: JN (wft)
 CP (SM 1)