



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: Alexander

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB110899

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: pH MeterTempProbe

Reagent/Standard	Lot/Log #
conductivity standard	WP86113

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB110899BL		1	25.0	0.77	0.77	09/10/2020	12:40
2	LB110899BS	1413	1	25.2	1405.00	1399.65	09/10/2020	12:42
3	L3959-01		1	25.1	74.60	74.46	09/10/2020	12:44
4	L3959-01DUP		1	25.1	73.90	73.76	09/10/2020	12:46

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

LB108999

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Conductance-91020 WorkList ID : 142203 Department : Wet-Chemistry Date : 09-10-2020 12:20:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2020	Water	L3959-01	A	Cool 4 deg C	STER08	L12	POOL-WATER	09/09/2020	SM2510 B

Date/Time 9/10/2020 12:30
 Received by: AM(wetchem)
 Relinquished by: JA (solt) Orlm

Date/Time 9/10/2020 13:10
 Received by: JA (wet) Orlm
 Relinquished by: AM(wetchem)