


LB110 901

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5040e-004

C= 2.0429e-002

Rho= 0.9999397

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.030	0.030	37	0.000	37					
0.05	0.050					159					
0.20	0.200	0.224	0.024	814	0.0 %	814					
2.50	2.500	2.483	-0.017	9836	0.0 %	9836					
5.00	5.000	4.931	-0.069	19611	0.0 %	19611					
7.50	7.500	7.478	-0.022	29784	0.0 %	29784					
10.0	10.000	10.054	0.054	40069	0.0 %	40069					

%D
12
-1
-1
0
1

LB110901
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	37	-		0 ISM02.4	PPB	9/10/2020 15:09	Std
	0.2 S0.2	814	-		0.2 ISM02.4	PPB	9/10/2020 15:11	Std
	2.5 S2.5	9836	-		2.5 ISM02.4	PPB	9/10/2020 15:13	Std
	5 S5.0	19611	-		5 ISM02.4	PPB	9/10/2020 15:16	Std
	7.5 S7.5	29784	-		7.5 ISM02.4	PPB	9/10/2020 15:18	Std
	10 S10.0	40069	-		10 ISM02.4	PPB	9/10/2020 15:20	Std
ICV	ICV	14900	3.7513 -		ISM02.4	PPB	9/10/2020 15:23	SMPL
ICB	ICB	-134	-0.0131 -		ISM02.4	PPB	9/10/2020 15:25	SMPL
CCV087	CCV087	19778	4.9728 -		ISM02.4	PPB	9/10/2020 15:27	SMPL
CCB087	CCB087	-117	-0.0089 -		ISM02.4	PPB	9/10/2020 15:30	SMPL
PB131521BL	PBW521	-46	0.0089 -		ISM02.4	PPB	9/10/2020 15:32	SMPL
L3892-01	MEW290	2581	0.6667 -		ISM02.4	PPB	9/10/2020 15:34	SMPL
L3892-02	MEW289	-248	-0.0417 -		ISM02.4	PPB	9/10/2020 15:36	SMPL
L3892-03	MEW289D	-80	0.0004 -		ISM02.4	PPB	9/10/2020 15:39	SMPL
L3892-04	MEW289S	3823	0.9777 -		ISM02.4	PPB	9/10/2020 15:41	SMPL
L3892-05	MEW2A1	-107	-0.0064 -		ISM02.4	PPB	9/10/2020 15:43	SMPL
L3892-06	MEW294	137	0.0547 -		ISM02.4	PPB	9/10/2020 15:46	SMPL
L3892-07	MEW295	-124	-0.0106 -		ISM02.4	PPB	9/10/2020 15:48	SMPL
L3892-08	MEW2A0	-81	0.0001 -		ISM02.4	PPB	9/10/2020 15:50	SMPL
CCV088	CCV088	19428	4.8851 -		ISM02.4	PPB	9/10/2020 15:52	SMPL
CCB088	CCB088	-50	0.0079 -		ISM02.4	PPB	9/10/2020 15:55	SMPL