



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110903

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	09/10/2020	15:10
2	CAL2	1	Water	NA	NA	20.5	7.00	09/10/2020	15:11
3	CAL3	1	Water	NA	NA	20.5	10.03	09/10/2020	15:12
4	ICV	1	Water	NA	NA	20.5	7.00	09/10/2020	15:13
5	CCV1	1	Water	NA	NA	20.5	2.01	09/10/2020	15:15
6	L3953-02	1	Water	NA	NA	23.8	6.48	09/10/2020	15:25
7	L3953-02DUP	1	Water	NA	NA	23.9	6.49	09/10/2020	15:26
8	CCV2	1	Water	NA	NA	20.4	12.02	09/10/2020	15:30

110903

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3953 WorkList ID : 142213 Department : Wet-Chemistry Date : 09-10-2020 15:18:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/10/2020	Water	L3953-02	-A Corrosivity	Cool 4 deg C	PSEG01	L12	N2930S	09/10/2020	9040C

Date/Time 09.10.2020 15:00
 Received by: JP (wet chem)
 Relinquished by: KS (wet chem)

Date/Time 09.10.2020 16:13
 Received by: KS (wet chem)
 Relinquished by: JP (wet chem)