



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110914

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	09/10/2020	17:35
2	CAL2	1	Water	NA	NA	20.1	7.00	09/10/2020	17:36
3	CAL3	1	Water	NA	NA	20.1	10.03	09/10/2020	17:38
4	ICV	1	Water	NA	NA	20.2	6.99	09/10/2020	17:39
5	CCV1	1	Water	NA	NA	20.2	2.01	09/10/2020	17:40
6	L3941-01	1	Solid	20.03	20	26.7	8.62	09/10/2020	17:42
7	L3976-01	1	Solid	20.04	20	25.1	6.02	09/10/2020	17:43
8	L3976-02	1	Solid	20.02	20	25.5	7.60	09/10/2020	17:45
9	L3976-02DUP	1	Solid	20.01	20	25.6	7.61	09/10/2020	17:46
10	CCV2	1	Water	NA	NA	20.2	12.02	09/10/2020	17:49

18109114

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3976 WorkList ID : 142218 Department : Wet-Chemistry Date : 09-10-2020 16:58:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/09/2020	Solid	L3941-01	Corrosivity	Cool 4 deg C	PSEG01	M41	DRUM-1	09/09/2020	9045D
09/09/2020	Solid	L3976-01	Corrosivity	Cool 4 deg C	PARS15	P12	BSY-WC-S-SB-0102	09/09/2020	9045D
09/09/2020	Solid	L3976-02	Corrosivity	Cool 4 deg C	PARS15	P12	BSY-WC-S-SB-03	09/09/2020	9045D

Date/Time 09/10/2020 17:10:00
 Received by: JO (wetchem)
 Relinquished by: KS (wetchem)

Date/Time 09/10/2020 18:20
 Received by: KS (wetchem)
 Relinquished by: JO (wetchem) *mm*