



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110917

Slope : 99.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/11/2020	13:35
2	CAL2	1	Water	NA	NA	20.1	7.00	09/11/2020	13:36
3	CAL3	1	Water	NA	NA	20.1	10.02	09/11/2020	13:37
4	ICV	1	Water	NA	NA	20.2	7.00	09/11/2020	13:38
5	CCV1	1	Water	NA	NA	20.2	2.02	09/11/2020	13:39
6	L3974-20	1	Water	NA	NA	23.9	9.08	09/11/2020	13:40
7	L3985-01	1	Water	NA	NA	22.7	4.48	09/11/2020	13:41
8	L3986-01	1	Water	NA	NA	22.6	7.64	09/11/2020	13:42
9	L3986-01DUP	1	Water	NA	NA	22.7	7.65	09/11/2020	13:43
10	CCV2	1	Water	NA	NA	20.2	12.02	09/11/2020	13:45

WORKLIST(Hardcopy Internal Chain)

287 110917

WorkList Name : PH L3985 WorkList ID : 142255 Department : Wet-Chemistry Date : 09-11-2020 13:27:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/07/2020	Water	L3985-01	pH	Cool 4 deg C	STER15	P12	GILROY-OMEGA-CELL	09/07/2020	SM 4500-H B
09/10/2020	Water	L3986-01	pH	Cool 4 deg C	STER08	P12	POOL-WATER	09/10/2020	SM 4500-H B

Date/Time 09-11-2020 13:30
 Received by: JS (wet chem)
 Relinquished by: JS (wet chem)

Date/Time 09-11-2020 14:40
 Received by: KS (wet chem)
 Relinquished by: JP (wet chem)