



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110918

Slope : 99.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/11/2020	13:35
2	CAL2	1	Water	NA	NA	20.1	7.00	09/11/2020	13:36
3	CAL3	1	Water	NA	NA	20.1	10.02	09/11/2020	13:37
4	ICV	1	Water	NA	NA	20.2	7.00	09/11/2020	13:38
5	CCV1	1	Water	NA	NA	20.2	2.02	09/11/2020	13:39
6	L3974-21	1	Water	NA	NA	23.9	9.08	09/11/2020	13:40
7	L3976-03	1	Water	NA	NA	20.3	7.12	09/11/2020	13:44
8	L3976-03DUP	1	Water	NA	NA	20.4	7.13	09/11/2020	13:45
9	CCV2	1	Water	NA	NA	20.2	12.02	09/11/2020	13:46

110918

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3986 WorkList ID : 142256 Department : Wet-Chemistry Date : 09-11-2020 13:28:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/09/2020	Water	L3976-03	pH	Cool 4 deg C	PARS15	P12	BSY-WC-WATER-DRUM	09/09/2020	9040C

Date/Time 09-11-2020 13:30
Received by: JD (wet-Ch(m))
Relinquished by: KS (wet-Ch(m))

Date/Time 09-11-2020 14:40
Received by: KS (wet-Ch(m))
Relinquished by: JD (wet-Ch(m))