



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110924

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/11/2020	14:40
2	CAL2	1	Water	NA	NA	20.3	7.00	09/11/2020	14:41
3	CAL3	1	Water	NA	NA	20.3	10.03	09/11/2020	14:42
4	ICV	1	Water	NA	NA	20.3	7.01	09/11/2020	14:44
5	CCV1	1	Water	NA	NA	20.3	2.02	09/11/2020	14:45
6	L3979-01	1	Water	NA	NA	25.1	7.47	09/11/2020	14:46
7	L3979-02	1	Water	NA	NA	25.6	7.09	09/11/2020	14:47
8	L3979-03	1	Water	NA	NA	25.5	6.89	09/11/2020	14:49
9	L3979-06	1	Water	NA	NA	25.3	6.84	09/11/2020	14:50
10	L3980-01	1	Water	NA	NA	25.1	5.03	09/11/2020	14:52
11	L3987-01	1	Water	NA	NA	19.9	7.74	09/11/2020	15:00
12	L3987-01DUP	1	Water	NA	NA	20.0	7.75	09/11/2020	15:01
13	L3994-01	1	Water	NA	NA	22.1	5.28	09/11/2020	15:03
14	CCV2	1	Water	NA	NA	20.2	12.02	09/11/2020	15:05

7B 110924

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3979 WorkList ID : 142259 Department : Wet-Chemistry Date : 09-11-2020 14:32:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2020	Water	L3979-01	pH	Cool 4 deg C	PSEG01	L12	31240	09/11/2020	9040C
09/11/2020	Water	L3979-02	pH	Cool 4 deg C	PSEG01	L12	31241	09/11/2020	9040C
09/11/2020	Water	L3979-03	pH	Cool 4 deg C	PSEG01	L12	31242	09/11/2020	9040C
09/11/2020	Water	L3979-06	pH	Cool 4 deg C	PSEG01	L12	31243	09/11/2020	9040C
09/11/2020	Water	L3980-01	pH	Cool 4 deg C	PSEG01	P21	31239	09/11/2020	9040C
09/10/2020	Water	L3987-01	pH	Cool 4 deg C	MINO01	P13	012-OUTFALL	09/10/2020	9040C
09/11/2020	Water	L3994-01	pH	Cool 4 deg C	MAJO01	P13	DRAIN-WATER-TANK-1	09/11/2020	9040C

Date/Time 09-11-2020 14:35
 Received by: JD
 Relinquished by: JD

Date/Time 09-11-2020 16:10
 Received by: JD
 Relinquished by: JD