



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110925

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP86746
Turbidity Calibration std, 20NTU	WP86742
Turbidity Calibration std, 40NTU	WP86741
Turbidity Calibration std, 1NTU	WP86745
Turbidity Calibration std, 5NTU	WP86744
Turbidity Calibration - CCV std, 10 NTU	WP86743
10 NTU Standard 500 ml	W2489

Intercept: 0.2815

Slope: 0.9627

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.282	09/11/2020	16:00
2	CAL2	1	1	1.180	09/11/2020	15:52
3	CAL3	5	1	5.210	09/11/2020	15:54
4	CAL4	10	1	9.880	09/11/2020	15:56
5	CAL5	20	1	19.500	09/11/2020	15:57
6	CAL6	40	1	38.800	09/11/2020	15:59



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB110925

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	10.407	09/11/2020	16:01
2	ICB		1	0.258	-0.024	09/11/2020	16:03
3	CCV1	10	1	9.920	10.012	09/11/2020	16:05
4	CCB1		1	0.259	-0.023	09/11/2020	16:07
5	LB110925BL		1	0.282	0.001	09/11/2020	16:08
6	L3921-01		1	33.200	34.194	09/11/2020	16:10
7	L3921-01DUP		1	33.600	34.609	09/11/2020	16:12
8	CCV2	10	1	9.840	9.929	09/11/2020	16:14
9	CCB2		1	0.269	-0.013	09/11/2020	16:16

LB110995 WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-91120

Worklist ID : 142264

Department : Wet-Chemistry

Date : 09-11-2020 15:34:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2020	Water	L3921-01	Turbidity	Cool 4 deg C	LIR001	M31	UDS-4-26	09/03/2020	SM2130 B

Date/Time

9/11/2020 15:40

Received by:

AMGwertchem

Relinquished by:

AMGwertchem

Date/Time

9/11/2020 16:20

Received by:

AMGwertchem

Relinquished by:

AMGwertchem