



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110929

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/11/2020	14:40
2	L3952-11		1	8	55.00	131.0	131.0	09/11/2020	15:10
3	L3952-11DUP		1	8	57.00	134.6	134.6	09/11/2020	15:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB110929

WorkList Name : fp91120-1 WorkList ID : 142247 Department : Wet-Chemistry Date : 09-11-2020 09:08:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2020	Solid	L3952-11 B	Flash Point	Cool 4 deg C	PSEG01	P11	16745	09/10/2020	1010A

Date/Time 09/11/2020 14:15
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 09/11/2020 15:40
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)