



Analytical Summary Report

Analysis Method: 9050A

ANALYST: Alexander

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB110938

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: pH MeterTempProbe

Reagent/Standard	Lot/Log #
conductivity standard	WP86113

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB110938BL		1	24.4	0.78	0.79	09/11/2020	12:30
2	LB110938BS	1413	1	25.1	1398.00	1395.33	09/11/2020	12:32
3	L3974-09		1	24.5	725.00	731.99	09/11/2020	12:42

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$