



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110939

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/11/2020	14:40
2	CAL2	1	Water	NA	NA	20.3	7.00	09/11/2020	14:41
3	CAL3	1	Water	NA	NA	20.3	10.03	09/11/2020	14:42
4	ICV	1	Water	NA	NA	20.3	7.01	09/11/2020	14:44
5	CCV1	1	Water	NA	NA	20.3	2.02	09/11/2020	14:45
6	L3969-04	1	Water	NA	NA	22.1	5.28	09/11/2020	15:03
7	L3969-04DUP	1	Water	NA	NA	22.2	5.29	09/11/2020	15:04
8	CCV2	1	Water	NA	NA	20.2	12.02	09/11/2020	15:05

18 110939

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14994 WorkList ID : 142303 Department : Wet-Chemistry Date : 09-11-2020 14:32:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2020	Water	L3969-04	pH	Cool 4 deg C	MAJO01	P12	DRAIN-WATER-TANK-1	09/11/2020	9040C

Date/Time 09.11.2020 14:35
Received by: JA W B - GEM
Relinquished by: JA

Date/Time 09.11.2020 16:10
Received by: JA
Relinquished by: JA