

LB 110940


INSTRUMENT ID: CV1

7470A

Linear

 μ Abs.:

37405



A= 0.0000e+000

B= 2.6176e-004

C= 2.7546e-002

Rho= 0.9992556

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	-45	0.000	-45				
0.2	0.200	0.183	-0.017	595	0.0 %	595				
2.5	2.500	2.472	-0.028	9339	0.0 %	9339				
5.0	5.000	4.925	-0.075	18708	0.0 %	18708				
7.5	7.500	7.786	0.286	29639	0.0 %	29639				
10.0	10.000	9.819	-0.181	37405	0.0 %	37405				

LB110940 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 14 Sep 2020 11:12:44					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	-45	-	0.0000	7470A	PPB	14 Sep 2020 11:25:48	S	Std
0.2 - 1	50.2	595	-	0.2000	7470A	PPB	14 Sep 2020 11:28:05	S	Std
2.5 - 1	52.5	9339	-	2.5000	7470A	PPB	14 Sep 2020 11:30:22	S	Std
5.0 - 1	55	18708	-	5.0000	7470A	PPB	14 Sep 2020 11:32:39	S	Std
7.5 - 1	57.5	29639	-	7.5000	7470A	PPB	14 Sep 2020 11:34:55	S	Std
10.0 - 1	510	37405	-	10.0000	7470A	PPB	14 Sep 2020 11:37:11	S	Std
ICV11 - 1	ICV11	14012	3.6953	-	7470A	PPB	14 Sep 2020 11:39:55	U	SMPL
ICB11 - 1	ICB11	-269	-0.0429	-	7470A	PPB	14 Sep 2020 11:42:10	U	SMPL
CCV01 - 1	CCV01	18791	4.9462	-	7470A	PPB	14 Sep 2020 11:44:24	U	SMPL
CCB01 - 1	CCB01	-264	-0.0416	-	7470A	PPB	14 Sep 2020 11:46:39	U	SMPL
CRA - 1	CRA	566	0.1757	-	7470A	PPB	14 Sep 2020 11:48:55	U	SMPL
HighStd - 1	HighStd	37814	9.9256	-	7470A	PPB	14 Sep 2020 11:51:10	U	SMPL
ChkStd - 1	ChkStd	26887	7.0654	-	7470A	PPB	14 Sep 2020 11:53:26	U	SMPL
PB131548BL - 1	PBW	-371	-0.0696	-	7470A	PPB	14 Sep 2020 11:55:44	U	SMPL
PB131548BS - 1	LCSW	14069	3.7102	-	7470A	PPB	14 Sep 2020 11:58:01	U	SMPL
L3953-02 - 1	N2930S	-260	-0.0405	-	7470A	PPB	14 Sep 2020 12:00:17	U	SMPL
L3953-02DUP - 1	N2930SDUP	-115	-0.0026	-	7470A	PPB	14 Sep 2020 12:02:34	U	SMPL
L3953-02MS - 1	N2930SMS	14759	3.8908	-	7470A	PPB	14 Sep 2020 12:04:51	U	SMPL
L3953-02MSD - 1	N2930SMSD	14336	3.7801	-	7470A	PPB	14 Sep 2020 12:07:08	U	SMPL
PB131549BL - 1	PBW	-253	-0.0387	-	7470A	PPB	14 Sep 2020 12:09:23	U	SMPL
CCV02 - 1	CCV02	18987	4.9975	-	7470A	PPB	14 Sep 2020 12:11:39	U	SMPL
CCB02 - 1	CCB02	-234	-0.0337	-	7470A	PPB	14 Sep 2020 12:13:54	U	SMPL
PB131549BS - 1	LCSW	13784	3.6356	-	7470A	PPB	14 Sep 2020 12:16:10	U	SMPL
L3952-01 - 1	40557	2466	0.6730	-	7470A	PPB	14 Sep 2020 12:18:25	U	SMPL
L3952-05 - 1	16744	234	0.0888	-	7470A	PPB	14 Sep 2020 12:20:41	U	SMPL
L3953-01 - 1	N2930S	0	0.0275	-	7470A	PPB	14 Sep 2020 12:22:59	U	SMPL
L3953-01DUP - 1	N2930SDUP	97	0.0529	-	7470A	PPB	14 Sep 2020 12:25:16	U	SMPL
L3953-01MS - 1	N2930SMS	16820	4.4303	-	7470A	PPB	14 Sep 2020 12:27:32	U	SMPL
L3953-01MSD - 1	N2930SMSD	15508	4.0869	-	7470A	PPB	14 Sep 2020 12:29:49	U	SMPL
L3976-03 - 1	BSY-WC-WATER-DRUM	13374	3.5283	-	7470A	PPB	14 Sep 2020 12:32:06	U	SMPL
PB131554BL - 1	PBW	-352	-0.0646	-	7470A	PPB	14 Sep 2020 12:34:23	U	SMPL
PB131554BS - 1	LCSW	17050	4.4905	-	7470A	PPB	14 Sep 2020 12:36:47	U	SMPL
CCV03 - 1	CCV03	19094	5.0255	-	7470A	PPB	14 Sep 2020 12:39:02	U	SMPL
CCB03 - 1	CCB03	-154	-0.0128	-	7470A	PPB	14 Sep 2020 12:41:18	U	SMPL
L3976-01 - 1	BSY-WC-S-SB-0102	480	0.1532	-	7470A	PPB	14 Sep 2020 12:43:34	U	SMPL
L3976-01DUP - 1	BSY-WC-S-SB-0102DUP	449	0.1451	-	7470A	PPB	14 Sep 2020 12:48:51	U	SMPL
L3976-01MS - 1	BSY-WC-S-SB-0102MS	14110	3.7209	-	7470A	PPB	14 Sep 2020 12:51:06	U	SMPL
L3976-01MSD - 1	BSY-WC-S-SB-0102MSD	13975	3.6856	-	7470A	PPB	14 Sep 2020 12:53:22	U	SMPL
L3976-02 - 1	BSY-WC-S-SB-03	-29	0.0200	-	7470A	PPB	14 Sep 2020 12:55:38	U	SMPL
PB131548TB - 1	PB131533TB	38	0.0375	-	7470A	PPB	14 Sep 2020 13:02:50	U	SMPL
PB131554TB - 1	PB131534TB	29	0.0351	-	7470A	PPB	14 Sep 2020 13:05:06	U	SMPL
CCV04 - 1	CCV04	18665	4.9132	-	7470A	PPB	14 Sep 2020 13:07:23	U	SMPL
CCB04 - 1	CCB04	-158	-0.0138	-	7470A	PPB	14 Sep 2020 13:09:41	U	SMPL
L3953-02LX5 - 1		-72	0.0087	-	7470A	PPB	14 Sep 2020 13:11:57	U	SMPL
L3953-02A - 1		14226	3.7513	-	7470A	PPB	14 Sep 2020 13:14:12	U	SMPL
L3953-01LX5 - 1		-134	-0.0075	-	7470A	PPB	14 Sep 2020 13:16:28	U	SMPL
L3953-01A - 1		14188	3.7414	-	7470A	PPB	14 Sep 2020 13:18:44	U	SMPL
L3976-01LX5 - 1		-39	0.0173	-	7470A	PPB	14 Sep 2020 13:20:59	U	SMPL
L3976-01A - 1		15422	4.0644	-	7470A	PPB	14 Sep 2020 13:23:15	U	SMPL
CCV05 - 1	CCV05	18847	4.9609	-	7470A	PPB	14 Sep 2020 13:25:30	U	SMPL
CCB05 - 1	CCB05	-236	-0.0342	-	7470A	PPB	14 Sep 2020 13:27:46	U	SMPL