


23110946

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.029	-0.029	-19	0.000	-19				
0.2	0.200	0.195	-0.005	1226	0.0 %	1226				
2.5	2.500	2.513	0.013	14125	0.0 %	14125				
5.0	5.000	4.982	-0.018	27858	0.0 %	27858				
7.5	7.500	7.643	0.143	42665	0.0 %	42665				
10.0	10.000	9.896	-0.104	55200	0.0 %	55200				

LB110946 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 14 Sep 2020 14:17:01				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	-19	-	0.0000	7471B	PPB	14 Sep 2020 14:19:29	S Std
0.2 - 1	50.2	1226	-	0.2000	7471B	PPB	14 Sep 2020 14:21:46	S Std
2.5 - 1	52.5	14125	-	2.5000	7471B	PPB	14 Sep 2020 14:24:03	S Std
5.0 - 1	55	27858	-	5.0000	7471B	PPB	14 Sep 2020 14:26:19	S Std
7.5 - 1	57.5	42665	-	7.5000	7471B	PPB	14 Sep 2020 14:28:36	S Std
10.0 - 1	59	55200	-	10.0000	7471B	PPB	14 Sep 2020 14:30:53	S Std
ICV12 - 1	ICV12	22480	4.0150	-	7471B	PPB	14 Sep 2020 14:34:32	U SMPL
ICB12 - 1	ICB12	-160	-0.0539	-	7471B	PPB	14 Sep 2020 14:36:47	U SMPL
CCV06 - 1	CCV06	28092	5.0236	-	7471B	PPB	14 Sep 2020 14:39:04	U SMPL
CCB06 - 1	CCB06	-149	-0.0519	-	7471B	PPB	14 Sep 2020 14:41:19	U SMPL
CRA - 1	CRA	1185	0.1878	-	7471B	PPB	14 Sep 2020 14:43:35	U SMPL
HighStd - 1	HighStd	54288	9.7317	-	7471B	PPB	14 Sep 2020 14:45:50	U SMPL
ChkStd - 1	ChkStd	39680	7.1063	-	7471B	PPB	14 Sep 2020 14:48:06	U SMPL
PB131577BL - 1	PB131577BL	-189	-0.0591	-	7471B	PPB	14 Sep 2020 14:50:24	U SMPL
PB131577BS - 1	PB131577BS	20111	3.5893	-	7471B	PPB	14 Sep 2020 14:52:41	U SMPL
L3863-01 - 1	HD-02-091120	6054	1.0629	-	7471B	PPB	14 Sep 2020 14:54:57	U SMPL
L3869-01 - 1	OR-03-091120	3915	0.6785	-	7471B	PPB	14 Sep 2020 14:57:15	U SMPL
L3981-01 - 1	B17-0-2	59143	10.6043	-	7471B	PPB	14 Sep 2020 14:59:32	U SMPL
L3981-02 - 1	B17-4-6	41838	7.4941	-	7471B	PPB	14 Sep 2020 15:01:50	U SMPL
L3981-03 - 1	B17-7-9	150087	26.9491	-	7471B	PPB	14 Sep 2020 15:04:07	U SMPL
CCV07 - 1	CCV07	27329	4.8865	-	7471B	PPB	14 Sep 2020 15:06:24	U SMPL
CCB07 - 1	CCB07	-192	-0.0597	-	7471B	PPB	14 Sep 2020 15:08:43	U SMPL
L3981-04 - 1	B18-0-2	90944	16.3197	-	7471B	PPB	14 Sep 2020 15:11:00	U SMPL
L3981-05 - 1	B18-4-6	16607	2.9595	-	7471B	PPB	14 Sep 2020 15:13:15	U SMPL
L3981-06 - 1	B18-7-9	48550	8.7004	-	7471B	PPB	14 Sep 2020 15:15:33	U SMPL
L3981-07 - 1	B19-0-2	219208	39.3718	-	7471B	PPB	14 Sep 2020 15:17:50	U SMPL
L3981-08 - 1	B19-4-6	54931	9.8473	-	7471B	PPB	14 Sep 2020 15:20:08	U SMPL
L3981-09 - 1	B19-7-9	46103	8.2607	-	7471B	PPB	14 Sep 2020 15:22:30	U SMPL
L3981-10 - 1	B20-0-2	29992	5.3651	-	7471B	PPB	14 Sep 2020 15:24:48	U SMPL
L3981-11 - 1	B20-4-6	32218	5.7652	-	7471B	PPB	14 Sep 2020 15:27:07	U SMPL
L3981-12 - 1	B20-7-9	8709	1.5401	-	7471B	PPB	14 Sep 2020 15:29:26	U SMPL
CCV08 - 1	CCV08	27921	4.9929	-	7471B	PPB	14 Sep 2020 15:31:44	U SMPL
CCB08 - 1	CCB08	-124	-0.0474	-	7471B	PPB	14 Sep 2020 15:33:59	U SMPL
L3993-03 - 1	PAV-B38-091120-0-10	24500	4.3781	-	7471B	PPB	14 Sep 2020 15:36:15	U SMPL
L3993-03DUP - 1	PAV-B38-091120-0-10DUP	24342	4.3497	-	7471B	PPB	14 Sep 2020 15:41:13	U SMPL
L3993-03MS - 1	PAV-B38-091120-0-10MS	49731	8.9127	-	7471B	PPB	14 Sep 2020 15:43:28	U SMPL
L3993-03MSD - 1	PAV-B38-091120-0-10MSD	48096	8.6188	-	7471B	PPB	14 Sep 2020 15:48:26	U SMPL
L3993-05 - 1	PAV-091120-DP	21578	3.8529	-	7471B	PPB	14 Sep 2020 15:50:41	U SMPL
L3993-09 - 1	PAV-B38-091120-10-17	6106	1.0722	-	7471B	PPB	14 Sep 2020 15:53:00	U SMPL
L3993-03LX5 - 1		4651	0.8107	-	7471B	PPB	14 Sep 2020 15:55:17	U SMPL
L3993-03A - 1		45239	8.1054	-	7471B	PPB	14 Sep 2020 15:57:33	U SMPL
CCV09 - 1	CCV09	27957	4.9994	-	7471B	PPB	14 Sep 2020 15:59:49	U SMPL
CCB09 - 1	CCB09	-123	-0.0473	-	7471B	PPB	14 Sep 2020 16:02:08	U SMPL
L3981-01DLX2 - 1		29880	5.3450	-	7471B	PPB	14 Sep 2020 16:04:23	U SMPL
L3981-03DLX5 - 1		29744	5.3205	-	7471B	PPB	14 Sep 2020 16:06:39	U SMPL
L3981-04DLX5 - 1		17848	3.1826	-	7471B	PPB	14 Sep 2020 16:08:56	U SMPL
L3981-07DLX10 - 1		23700	4.2343	-	7471B	PPB	14 Sep 2020 16:11:12	U SMPL
CCV10 - 1	CCV10	27879	4.9854	-	7471B	PPB	14 Sep 2020 16:16:03	U SMPL
CCB10 - 1	CCB10	-112	-0.0453	-	7471B	PPB	14 Sep 2020 16:18:18	U SMPL