



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110953

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/14/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	09/14/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	09/14/2020	17:12
4	ICV	1	Water	NA	NA	20.3	7.00	09/14/2020	17:15
5	CCV1	1	Water	NA	NA	20.2	2.01	09/14/2020	17:16
6	L3969-05	1	Water	NA	NA	18.6	6.03	09/14/2020	17:40
7	L3969-05DUP	1	Water	NA	NA	18.7	6.04	09/14/2020	17:41
8	L4008-02	1	Water	NA	NA	24.9	6.14	09/14/2020	17:42
9	CCV2	1	Water	NA	NA	20.3	12.03	09/14/2020	17:45

MP 110953

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3869 WorkList ID : 142313 Department : Wet-Chemistry Date : 09-14-2020 13:22:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/14/2020	Water	L3969-05	pH	Cool 4 deg C	MAJO01	P12	DRAIN-WATER-TANK-1	09/14/2020	9040C
09/14/2020	Water	L4008-02	pH	Cool 4 deg C	PSEG01	P33	13703	09/14/2020	9040C

Date/Time 09.16.2020 16:11:00
Received by: Jn (wet-chem)
Relinquished by: KS (wet-chem)

Date/Time 09.16.2020 17:25
Received by: KS (wet-chem)
Relinquished by: Jn (wet-chem)