



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110954

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/14/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	09/14/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	09/14/2020	17:12
4	ICV	1	Water	NA	NA	20.3	7.00	09/14/2020	17:15
5	CCV1	1	Water	NA	NA	20.2	2.01	09/14/2020	17:16
6	L4005-05	1	Solid	20.03	20	24.7	7.10	09/14/2020	17:45
7	L4005-05DUP	1	Solid	20.04	20	24.8	7.11	09/14/2020	17:46
8	L4005-06	1	Solid	20.03	20	24.6	8.81	09/14/2020	17:49
9	L4008-01	1	Solid	20.02	20	25.7	7.20	09/14/2020	17:50
10	CCV2	1	Water	NA	NA	20.2	12.03	09/14/2020	17:51

WORKLIST(Hardcopy Internal Chain)

110954

WorkList Name : PH L4005 S

WorkList ID : 142315

Department : Wet-Chemistry

Date : 09-14-2020 16:27:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/14/2020	Solid	L4005-05	pH	Cool 4 deg C	PSEG01	P33	35-37	09/14/2020	9045D
09/14/2020	Solid	L4005-06	pH	Cool 4 deg C	PSEG01	P33	32621	09/14/2020	9045D
09/14/2020	Solid	L4008-01	pH	Cool 4 deg C	PSEG01	P33	40951	09/14/2020	9045D

Date/Time 09/14/2020 16:00
 Received by: Jn (wetchem)
 Relinquished by: KS (wetchem)

Date/Time 09/14/2020 17:35
 Received by: KS (wetchem)
 Relinquished by: Jn (wetchem)