



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110956

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/14/2020	09:45
2	L3979-04		1	8	100.00	>212.0	>212.0	09/14/2020	10:15
3	L3979-04DUP		1	8	100.00	>212.0	>212.0	09/14/2020	10:40
4	L3979-06		1	9	100.00	>212.0	>212.0	09/14/2020	11:10
5	L3980-01		1	8	100.00	>212.0	>212.0	09/14/2020	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp91420 WorkList ID : 142299 Department : Wet-Chemistry Date : 09-14-2020 09:17:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/16/2020	Water	L3979-04	Flash Point	Cool 4 deg C	PSEG01	L12	40-42	09/11/2020	1010A
09/25/2020	Water	L3979-06	Flash Point	Cool 4 deg C	PSEG01	L12	31243	09/11/2020	1010A
09/25/2020	Water	L3980-01	Flash Point	Cool 4 deg C	PSEG01	P21	31239	09/11/2020	1010A

Date/Time 09/14/2020 09:25
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 09/14/2020 11:53
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)