


LB110966

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	-30	0.000	-30				
0.2	0.200	0.223	0.023	671	0.0 %	671				
2.5	2.500	2.483	-0.017	9018	0.0 %	9018				
5.0	5.000	5.037	0.037	18456	0.0 %	18456				
7.5	7.500	7.283	-0.217	26751	0.0 %	26751				
10.0	10.000	10.141	0.141	37310	0.0 %	37310				

LB110966 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Sep 2020 09:42:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-30	-	0.00007470A	PPB		15 Sep 2020 09:47:26	S	Std
0.2 - 1	50.2	671	-	0.20007470A	PPB		15 Sep 2020 09:49:43	S	Std
2.5 - 1	52.5	9018	-	2.50007470A	PPB		15 Sep 2020 09:52:00	S	Std
5.0 - 1	55	18456	-	5.00007470A	PPB		15 Sep 2020 09:54:16	S	Std
7.5 - 1	57.5	26751	-	7.50007470A	PPB		15 Sep 2020 09:56:33	S	Std
10.0 - 1	510	37310	-	10.00007470A	PPB		15 Sep 2020 09:58:49	S	Std
ICV13 - 1	ICV13	13302	3.6422	-7470A	PPB		15 Sep 2020 10:04:26	U	SMPL
ICB13 - 1	ICB13	-230	-0.0209	-7470A	PPB		15 Sep 2020 10:06:41	U	SMPL
CCV11 - 1	CCV11	18604	5.0774	-7470A	PPB		15 Sep 2020 10:08:57	U	SMPL
CCB11 - 1	CCB11	-250	-0.0263	-7470A	PPB		15 Sep 2020 10:11:12	U	SMPL
CRA - 1	CRA	627	0.2111	-7470A	PPB		15 Sep 2020 10:13:28	U	SMPL
HighStd - 1	HighStd	36955	10.0450	-7470A	PPB		15 Sep 2020 10:15:43	U	SMPL
ChkStd - 1	ChkStd	23745	6.4691	-7470A	PPB		15 Sep 2020 10:17:58	U	SMPL
PB131599BL - 1	PBW	-147	0.0016	-7470A	PPB		15 Sep 2020 10:20:16	U	SMPL
PB131599BS - 1	LCSW	13318	3.6465	-7470A	PPB		15 Sep 2020 10:22:33	U	SMPL
L3863-04 - 1	HD-02-091120	341	0.1337	-7470A	PPB		15 Sep 2020 10:24:50	U	SMPL
L3863-04DUP - 1	HD-02-091120DUP	268	0.1139	-7470A	PPB		15 Sep 2020 10:27:06	U	SMPL
L3863-04MS - 1	HD-02-091120MS	14967	4.0929	-7470A	PPB		15 Sep 2020 10:29:23	U	SMPL
L3863-04MSD - 1	HD-02-091120MSD	14040	3.8420	-7470A	PPB		15 Sep 2020 10:31:40	U	SMPL
L3869-05 - 1	OR-03-091120	-231	-0.0212	-7470A	PPB		15 Sep 2020 10:33:56	U	SMPL
CCV12 - 1	CCV12	19059	5.2006	-7470A	PPB		15 Sep 2020 10:36:12	U	SMPL
CCB12 - 1	CCB12	-221	-0.0185	-7470A	PPB		15 Sep 2020 10:38:27	U	SMPL
L3981-13 - 1	B18-0-2	-41	0.0303	-7470A	PPB		15 Sep 2020 10:40:45	U	SMPL
L3981-14 - 1	B18-4-6	23	0.0476	-7470A	PPB		15 Sep 2020 10:43:00	U	SMPL
L3981-15 - 1	B18-7-9	64	0.0587	-7470A	PPB		15 Sep 2020 10:45:16	U	SMPL
L3981-16 - 1	B19-0-2	125	0.0752	-7470A	PPB		15 Sep 2020 10:47:32	U	SMPL
L3981-17 - 1	B19-4-6	-13	0.0378	-7470A	PPB		15 Sep 2020 10:49:48	U	SMPL
L3981-18 - 1	B19-7-9	7	0.0433	-7470A	PPB		15 Sep 2020 10:52:05	U	SMPL
L3981-19 - 1	B20-0-2	18	0.0462	-7470A	PPB		15 Sep 2020 10:54:22	U	SMPL
L3981-20 - 1	B20-4-6	36	0.0511	-7470A	PPB		15 Sep 2020 10:56:39	U	SMPL
L3981-21 - 1	B20-7-9	48	0.0544	-7470A	PPB		15 Sep 2020 10:58:56	U	SMPL
L3993-04 - 1	PAV-B38-091120-0-10	88	0.0652	-7470A	PPB		15 Sep 2020 11:01:12	U	SMPL
CCV13 - 1	CCV13	18729	5.1113	-7470A	PPB		15 Sep 2020 11:03:27	U	SMPL
CCB13 - 1	CCB13	-193	-0.0109	-7470A	PPB		15 Sep 2020 11:05:43	U	SMPL
L3993-06 - 1	PAV-091120-DP	-36	0.0316	-7470A	PPB		15 Sep 2020 11:07:59	U	SMPL
L3993-10 - 1	PAV-B38-091120-10-17	-22	0.0354	-7470A	PPB		15 Sep 2020 11:10:15	U	SMPL
PB131600BL - 1	PBW	-17	0.0368	-7470A	PPB		15 Sep 2020 11:12:31	U	SMPL
PB131600BS - 1	LCSW	12765	3.4968	-7470A	PPB		15 Sep 2020 11:29:17	U	SMPL
L3993-13 - 1	FIELD-BLANK	-81	0.0194	-7470A	PPB		15 Sep 2020 11:31:32	U	SMPL
L3993-14 - 1	PAV-B41-091120-GW	4779	1.3350	-7470A	PPB		15 Sep 2020 11:33:48	U	SMPL
L3993-14DUP - 1	PAV-B41-091120-GWDUP	4783	1.3361	-7470A	PPB		15 Sep 2020 11:36:05	U	SMPL
L3993-14MS - 1	PAV-B41-091120-GWMS	19385	5.2888	-7470A	PPB		15 Sep 2020 11:38:21	U	SMPL
L3993-14MSD - 1	PAV-B41-091120-GWMSD	20684	5.6405	-7470A	PPB		15 Sep 2020 11:40:39	U	SMPL
L3993-20 - 1	PAV-B41-091120-GW	-90	0.0170	-7470A	PPB		15 Sep 2020 11:42:54	U	SMPL
CCV14 - 1	CCV14	18718	5.1083	-7470A	PPB		15 Sep 2020 11:45:10	U	SMPL
CCB14 - 1	CCB14	-41	0.0303	-7470A	PPB		15 Sep 2020 11:47:25	U	SMPL
PB131599TB - 1	PB131573TB	155	0.0833	-7470A	PPB		15 Sep 2020 11:52:20	U	SMPL
L3863-04LX5 - 1		141	0.0795	-7470A	PPB		15 Sep 2020 11:54:35	U	SMPL
L3863-04A - 1		13505	3.6971	-7470A	PPB		15 Sep 2020 11:56:51	U	SMPL
L3993-14LX5 - 1		1023	0.3183	-7470A	PPB		15 Sep 2020 11:59:06	U	SMPL
L3993-14A - 1		20290	5.5338	-7470A	PPB		15 Sep 2020 12:01:22	U	SMPL
CCV15 - 1	CCV15	18788	5.1272	-7470A	PPB		15 Sep 2020 12:03:39	U	SMPL
CCB15 - 1	CCB15	-44	0.0295	-7470A	PPB		15 Sep 2020 12:05:55	U	SMPL