



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110968

Slope : 97.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	09/15/2020	11:15
2	CAL2	1	Water	NA	NA	21.1	6.99	09/15/2020	11:16
3	CAL3	1	Water	NA	NA	21.1	10.03	09/15/2020	11:17
4	ICV	1	Water	NA	NA	21.1	7.00	09/15/2020	11:19
5	CCV1	1	Water	NA	NA	21.1	2.02	09/15/2020	11:20
6	L4023-01	1	Water	NA	NA	21.8	5.49	09/15/2020	11:22
7	L4023-01DUP	1	Water	NA	NA	21.9	5.50	09/15/2020	11:23
8	CCV2	1	Water	NA	NA	21.2	12.02	09/15/2020	11:25