



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110975

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	09/15/2020	13:10
2	CAL2	1	Water	NA	NA	21.2	7.00	09/15/2020	13:11
3	CAL3	1	Water	NA	NA	21.2	10.03	09/15/2020	13:12
4	ICV	1	Water	NA	NA	21.2	6.99	09/15/2020	13:15
5	CCV1	1	Water	NA	NA	21.2	2.01	09/15/2020	13:16
6	L3969-06	1	Water	NA	NA	18.4	6.12	09/15/2020	13:19
7	L3969-06DUP	1	Water	NA	NA	18.5	6.13	09/15/2020	13:20
8	CCV2	1	Water	NA	NA	20.1	12.02	09/15/2020	13:22

VB 110975

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3969-1 WorkList ID : 142349 Department : Wet-Chemistry Date : 09-15-2020 13:00:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/15/2020	Water	L3969-06	-pH	Cool 4 deg C	MAJO01	P12	DRAIN-WATER-TANK-1	09/15/2020	9040C

Date/Time 09-15-2020 13:05
Received by: JH/wet
Relinquished by: JH/wet

Date/Time 09-15-2020 14:15
Received by: JH/wet
Relinquished by: JH/wet