



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110989

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	09/16/2020	09:45
2	L4008-01		1	8	60.00	140.0	139.7	09/16/2020	10:15
3	L4008-01DUP		1	8	61.00	141.8	141.5	09/16/2020	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB110990

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp91520-1 WorkList ID : 142337 Department : Wet-Chemistry Date : 09-15-2020 08:56:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2020	Water	L4008-02	Flash Point	Cool 4 deg C	PSEG01	P33	13703	09/14/2020	1010A

Date/Time 09/16/2020 10:45
Received by: KS (wetchem)
Relinquished by: SQ (wetchem)

Date/Time 09/16/2020 12:00
Received by: JP (wetchem)
Relinquished by: KS (wetchem)