



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110991

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.00      | 80.6      | 80.3            | 09/16/2020 | 11:10     |
| 2   | L4008-02    |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 09/16/2020 | 11:40     |
| 3   | L4008-02DUP |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 09/16/2020 | 11:58     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp91520-1      WorkList ID : 142337      Department : Wet-Chemistry      Date : 09-15-2020 08:56:08

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 09/17/2020 | Water  | L4008-02 | Flash Point | Cool 4 deg C | PSEG01   | P33              | 13703           | 09/14/2020   | 1010A  |

LB 110991

Date/Time 09/16/2020 10:45  
 Received by: KS (wetchem)  
 Relinquished by: SQ (wetchem)

Date/Time 09/16/2020 12:00  
 Received by: JP (wetchem)  
 Relinquished by: KS (wetchem)